

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031225\
 Data File : VY021502.D
 Acq On : 12 Mar 2025 11:47
 Operator : SY/MD
 Sample : VY0312SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0312SBL01

Quant Time: Mar 13 01:36:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	279829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	501387	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.414	117	431961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	167294	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	158571	53.614	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.220%
35) Dibromofluoromethane	7.634	113	164760	49.993	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.980%
50) Toluene-d8	10.103	98	609608	48.853	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.700%
62) 4-Bromofluorobenzene	12.408	95	176618	41.609	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031225\
Data File : VY021502.D
Acq On : 12 Mar 2025 11:47
Operator : SY/MD
Sample : VY0312SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0312SBL01

Quant Time: Mar 13 01:36:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

