

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031225\
 Data File : VY021512.D
 Acq On : 12 Mar 2025 16:20
 Operator : SY/MD
 Sample : Q1524-01
 Misc : 5.06g/5.6mL/MSVOA_Y/SOIL/B
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11930-343-COMP

Quant Time: Mar 13 01:38:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	28826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	36326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	22171	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	4569	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	10535	34.578	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	69.160%
35) Dibromofluoromethane	7.628	113	11681	48.920	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.840%
50) Toluene-d8	10.109	98	40068	44.319	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.640%
62) 4-Bromofluorobenzene	12.408	95	6099	19.832	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	39.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031225\
Data File : VY021512.D
Acq On : 12 Mar 2025 16:20
Operator : SY/MD
Sample : Q1524-01
Misc : 5.06g/5.6mL/MSVOA_Y/SOIL/B
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11930-343-COMP

Quant Time: Mar 13 01:38:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

