

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031325\
 Data File : VY021525.D
 Acq On : 13 Mar 2025 15:08
 Operator : SY/MD
 Sample : Q1549-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11978

Quant Time: Mar 14 14:56:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	34466	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	52269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	41188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	10544	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	20627	56.623	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.240%
35) Dibromofluoromethane	7.634	113	20282	59.033	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.060%
50) Toluene-d8	10.109	98	61415	47.211	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.420%
62) 4-Bromofluorobenzene	12.402	95	13422	30.332	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	60.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031325\
Data File : VY021525.D
Acq On : 13 Mar 2025 15:08
Operator : SY/MD
Sample : Q1549-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11978

Quant Time: Mar 14 14:56:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

