

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031722\
 Data File : VY007888.D
 Acq On : 17 Mar 2022 20:36
 Operator : SY/MD
 Sample : N1875-11
 Misc : 1.30g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10C

Quant Time: Mar 18 08:57:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	7107	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	13284	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	12055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3472	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	5083	62.727	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.460%
35) Dibromofluoromethane	7.716	113	3732	40.003	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	80.000%
50) Toluene-d8	10.179	98	16030	57.826	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.660%
62) 4-Bromofluorobenzene	12.483	95	5163	38.561	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031722\
Data File : VY007888.D
Acq On : 17 Mar 2022 20:36
Operator : SY/MD
Sample : N1875-11
Misc : 1.30g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10C

Quant Time: Mar 18 08:57:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

