

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031722\
 Data File : VY007889.D
 Acq On : 17 Mar 2022 21:00
 Operator : SY/MD
 Sample : N1875-09
 Misc : 7.96g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10B

Quant Time: Mar 18 08:58:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	20828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	35286	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	29078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	8581	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	11054	46.547	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.100%
35) Dibromofluoromethane	7.715	113	10063	40.608	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.220%
50) Toluene-d8	10.178	98	41672	56.456	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.920%
62) 4-Bromofluorobenzene	12.483	95	11874	33.386	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	66.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031722\
Data File : VY007889.D
Acq On : 17 Mar 2022 21:00
Operator : SY/MD
Sample : N1875-09
Misc : 7.96g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10B

Quant Time: Mar 18 08:58:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

