



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031722\
 Data File : VY007867.D
 Acq On : 17 Mar 2022 09:08
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Title : SW846 8260
 Last Update : Thu Mar 17 13:55:34 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	23125	PASS
75	95	30	60	48.3	61960	PASS
95	95	100	100	100.0	128369	PASS
96	95	5	9	6.8	8767	PASS
173	174	0.00	2	0.8	815	PASS
174	95	50	100	76.8	98560	PASS
175	174	5	9	7.4	7300	PASS
176	174	95	101	96.2	94779	PASS
177	176	5	9	6.7	6303	PASS