

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031725\
 Data File : VY021549.D
 Acq On : 17 Mar 2025 14:26
 Operator : SY/MD
 Sample : Q1574-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC1

Quant Time: Mar 18 01:27:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	98896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	155221	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	116517	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	33476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	51252	49.032	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.060%
35) Dibromofluoromethane	7.634	113	54695	53.607	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.220%
50) Toluene-d8	10.109	98	177431	45.929	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.860%
62) 4-Bromofluorobenzene	12.408	95	39595	30.131	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	60.260%
Target Compounds						
7) Trichlorofluoromethane	3.050	101	2375	1.217	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031725\
Data File : VY021549.D
Acq On : 17 Mar 2025 14:26
Operator : SY/MD
Sample : Q1574-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC1

Quant Time: Mar 18 01:27:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

