

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031725\
 Data File : VY021551.D
 Acq On : 17 Mar 2025 15:32
 Operator : SY/MD
 Sample : Q1581-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-031425

Quant Time: Mar 18 01:27:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	3642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	5971	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	3963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	499	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	2950	76.636	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.280%
35) Dibromofluoromethane	7.622	113	2086	53.149	ug/l	-0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.300%
50) Toluene-d8	10.109	98	6537	43.989	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	87.980%
62) 4-Bromofluorobenzene	12.408	95	771	15.252	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	30.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031725\
Data File : VY021551.D
Acq On : 17 Mar 2025 15:32
Operator : SY/MD
Sample : Q1581-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-031425

Quant Time: Mar 18 01:27:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

