



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031725\
Data File : VY021542.D
Acq On : 17 Mar 2025 09:26
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Title : SW846 8260
Last Update : Wed Mar 05 12:42:45 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	31101	PASS
75	95	30	60	53.8	80184	PASS
95	95	100	100	100.0	148955	PASS
96	95	5	9	6.8	10182	PASS
173	174	0.00	2	1.2	1478	PASS
174	95	50	100	80.8	120405	PASS
175	174	5	9	7.9	9550	PASS
176	174	95	101	95.8	115325	PASS
177	176	5	9	6.7	7771	PASS