

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
 Data File : VY007897.D
 Acq On : 18 Mar 2022 13:13
 Operator : SY/MD
 Sample : N1919-02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Mar 19 00:12:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	301371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	544709	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	492203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	212235	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	179079	52.115	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.240%
35) Dibromofluoromethane	7.722	113	169588	44.331	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.660%
50) Toluene-d8	10.179	98	646670	56.786	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.580%
62) 4-Bromofluorobenzene	12.483	95	228223	41.569	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
Data File : VY007897.D
Acq On : 18 Mar 2022 13:13
Operator : SY/MD
Sample : N1919-02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Mar 19 00:12:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

