

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
 Data File : VY007902.D
 Acq On : 18 Mar 2022 15:10
 Operator : SY/MD
 Sample : N1945-08
 Misc : 6.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 4PH-1-7.5-8

Quant Time: Mar 19 01:36:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	281237	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	511153	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	467961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	199189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	171878	53.601	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.200%	
35) Dibromofluoromethane	7.716	113	162757	45.339	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 90.680%	
50) Toluene-d8	10.179	98	613701	57.502	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 115.000%	
62) 4-Bromofluorobenzene	12.483	95	214403	41.615	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 83.240%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
Data File : VY007902.D
Acq On : 18 Mar 2022 15:10
Operator : SY/MD
Sample : N1945-08
Misc : 6.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
4PH-1-7.5-8

Quant Time: Mar 19 01:36:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

