

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
 Data File : VY007906.D
 Acq On : 18 Mar 2022 16:44
 Operator : SY/MD
 Sample : N1951-08
 Misc : 5.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DB-13-(20-25)-3-12-22

Quant Time: Mar 19 01:37:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	265865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	483769	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	447704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	191693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	162312	53.544	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	107.080%	
35) Dibromofluoromethane	7.716	113	154367	45.436	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	90.880%	
50) Toluene-d8	10.179	98	581072	57.529	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	115.060%	
62) 4-Bromofluorobenzene	12.483	95	207614	42.579	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	85.160%	
Target Compounds						
16) Acetone	3.948	43	11425	14.207	ug/l	Qvalue # 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
Data File : VY007906.D
Acq On : 18 Mar 2022 16:44
Operator : SY/MD
Sample : N1951-08
Misc : 5.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DB-13-(20-25)-3-12-22

Quant Time: Mar 19 01:37:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

