

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
 Data File : VY007911.D
 Acq On : 18 Mar 2022 18:41
 Operator : SY/MD
 Sample : N1951-05
 Misc : 6.89g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DB-14-(13-15)-3-12-22

Quant Time: Mar 19 01:38:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	31895	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	59919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	50070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	13442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	23580	64.840	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 129.680%	
35) Dibromofluoromethane	7.716	113	19698	46.810	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 93.620%	
50) Toluene-d8	10.179	98	69003	54.892	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.780%	
62) 4-Bromofluorobenzene	12.477	95	20171	33.399	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 66.800%	
Target Compounds						
16) Acetone	3.942	43	6233	64.608	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
Data File : VY007911.D
Acq On : 18 Mar 2022 18:41
Operator : SY/MD
Sample : N1951-05
Misc : 6.89g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DB-14-(13-15)-3-12-22

Quant Time: Mar 19 01:38:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

