

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
 Data File : VY007929.D
 Acq On : 19 Mar 2022 03:13
 Operator : SY/MD
 Sample : N1958-09
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C-9-8-9

Quant Time: Mar 21 03:19:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	52911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	94901	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	83052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	32467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	32651	54.122	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.240%	
35) Dibromofluoromethane	7.716	113	29031	43.558	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 87.120%	
50) Toluene-d8	10.179	98	112800	56.862	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 113.720%	
62) 4-Bromofluorobenzene	12.483	95	36800	38.472	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 76.940%	
Target Compounds						
					Qvalue	
16) Acetone	3.942	43	3917	24.475	ug/l #	71
20) Methylene Chloride	4.710	84	6596	6.615	ug/l #	80
67) Ethyl Benzene	11.593	91	7477	2.110	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
Data File : VY007929.D
Acq On : 19 Mar 2022 03:13
Operator : SY/MD
Sample : N1958-09
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C-9-8-9

Quant Time: Mar 21 03:19:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

