

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
 Data File : VY007932.D
 Acq On : 19 Mar 2022 04:23
 Operator : SY/MD
 Sample : N1958-12
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C-12-11-12

Quant Time: Mar 21 03:20:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	245594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	458001	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	424260	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	184353	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	160952	57.478	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.960%
35) Dibromofluoromethane	7.716	113	144236	44.842	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.680%
50) Toluene-d8	10.179	98	566765	59.465	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	118.920%
62) 4-Bromofluorobenzene	12.483	95	200460	43.424	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.840%
Target Compounds						
16) Acetone	3.942	43	27743	37.346	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
Data File : VY007932.D
Acq On : 19 Mar 2022 04:23
Operator : SY/MD
Sample : N1958-12
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C-12-11-12

Quant Time: Mar 21 03:20:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

