

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
 Data File : VY007937.D
 Acq On : 19 Mar 2022 06:20
 Operator : SY/MD
 Sample : N1983-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 124035

Quant Time: Mar 21 03:21:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	63718	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	110567	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	95263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	45686	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	33753	46.459	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.920%
35) Dibromofluoromethane	7.715	113	32511	41.868	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.740%
50) Toluene-d8	10.178	98	131462	56.882	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.760%
62) 4-Bromofluorobenzene	12.483	95	46548	41.768	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
Data File : VY007937.D
Acq On : 19 Mar 2022 06:20
Operator : SY/MD
Sample : N1983-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
124035

Quant Time: Mar 21 03:21:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

