

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
 Data File : VY007938.D
 Acq On : 19 Mar 2022 06:44
 Operator : SY/MD
 Sample : N1983-03
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 124042

Quant Time: Mar 21 03:21:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	251562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	464693	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	421871	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	181339	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	162852	56.777	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.560%
35) Dibromofluoromethane	7.716	113	146497	44.889	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.780%
50) Toluene-d8	10.179	98	565799	58.405	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	116.800%
62) 4-Bromofluorobenzene	12.483	95	199484	42.591	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
Data File : VY007938.D
Acq On : 19 Mar 2022 06:44
Operator : SY/MD
Sample : N1983-03
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
124042

Quant Time: Mar 21 03:21:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

