

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
 Data File : VY007901.D
 Acq On : 18 Mar 2022 14:47
 Operator : SY/MD
 Sample : N1945-09
 Misc : 5.64g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 4PH-2-7.5-8

Quant Time: Mar 19 01:36:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	285828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	517753	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	468530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	202049	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	172908	53.056	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.120%
35) Dibromofluoromethane	7.716	113	163313	44.914	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.820%
50) Toluene-d8	10.179	98	619163	57.249	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.500%
62) 4-Bromofluorobenzene	12.483	95	213970	41.002	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.000%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031822\
Data File : VY007901.D
Acq On : 18 Mar 2022 14:47
Operator : SY/MD
Sample : N1945-09
Misc : 5.64g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
4PH-2-7.5-8

Quant Time: Mar 19 01:36:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY007901.D
Acq: 18 Mar 2022 14:47




Instrument :
MSVOA_Y
ClientSampleId :
4PH-2-7.5-8


Tgt Ion:168 Resp: 285828
Ion Ratio Lower Upper
168 100
99 59.3 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 53.056 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. 0.006 min
Lab File: VY007901.D
Acq: 18 Mar 2022 14:47

Tgt Ion: 65 Resp: 172908
Ion Ratio Lower Upper
65 100
67 52.6 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VY007901.D

Acq: 18 Mar 2022 14:47

Instrument :

MSVOA_Y

ClientSampleId :

4PH-2-7.5-8





Tgt Ion:114 Resp: 517753

Ion Ratio Lower Upper

114 100

63 18.6 0.0 43.2

88 15.3 0.0 31.8



#35

Dibromofluoromethane

Concen: 44.914 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY007901.D

Acq: 18 Mar 2022 14:47




Tgt Ion:113 Resp: 163313


Ion Ratio Lower Upper

113 100

111 103.1 82.6 124.0

192 17.8 17.6 26.4









#50
Toluene-d8
Concen: 57.249 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY007901.D
Acq: 18 Mar 2022 14:47




Instrument :
MSVOA_Y
ClientSampleId :
4PH-2-7.5-8


Tgt Ion: 98 Resp: 619163
Ion Ratio Lower Upper
98 100
100 65.7 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 41.002 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY007901.D
Acq: 18 Mar 2022 14:47

Tgt Ion: 95 Resp: 213970
Ion Ratio Lower Upper
95 100
174 77.2 0.0 177.0
176 75.8 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 117
Delta R.T. -0.000 min
Lab File: VY007901.D
Acq: 18 Mar 2022 14:47

Instrument :
MSVOA_Y
ClientSampleId :
4PH-2-7.5-8

Tgt Ion:117 Resp: 468530
Ion Ratio Lower Upper
117 100
82 54.7 43.0 64.4
119 31.4 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY007901.D
Acq: 18 Mar 2022 14:47

Tgt Ion:152 Resp: 202049
Ion Ratio Lower Upper
152 100
115 59.9 28.2 84.7
150 158.2 0.0 347.6

