

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031824\
 Data File : VY017692.D
 Acq On : 18 Mar 2024 16:37
 Operator : SY/MD
 Sample : P1764-02
 Misc : 4.42g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GRAB-2

Quant Time: Mar 19 04:10:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	112388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	211359	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	187810	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	69328	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	60245	56.561	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.120%
35) Dibromofluoromethane	7.722	113	69411	47.128	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.260%
50) Toluene-d8	10.185	98	242088	47.012	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.020%
62) 4-Bromofluorobenzene	12.489	95	64293	39.727	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.460%
Target Compounds						
20) Methylene Chloride	4.723	84	15472	8.163	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031824\
Data File : VY017692.D
Acq On : 18 Mar 2024 16:37
Operator : SY/MD
Sample : P1764-02
Misc : 4.42g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GRAB-2

Quant Time: Mar 19 04:10:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

