

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031824\
 Data File : VY017695.D
 Acq On : 18 Mar 2024 17:48
 Operator : SY/MD
 Sample : P1764-05
 Misc : 4.68g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GRAB-5

Quant Time: Mar 19 04:11:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	115069	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	213792	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	193037	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	73595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	62591	57.394	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.780%
35) Dibromofluoromethane	7.722	113	70817	47.536	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.080%
50) Toluene-d8	10.185	98	246497	47.323	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.640%
62) 4-Bromofluorobenzene	12.489	95	67152	41.022	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.040%
Target Compounds						
20) Methylene Chloride	4.716	84	15854	8.175	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031824\
 Data File : VY017695.D
 Acq On : 18 Mar 2024 17:48
 Operator : SY/MD
 Sample : P1764-05
 Misc : 4.68g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GRAB-5

Quant Time: Mar 19 04:11:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

