

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031824\
 Data File : VY017698.D
 Acq On : 18 Mar 2024 18:59
 Operator : SY/MD
 Sample : P1764-08
 Misc : 4.54g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GRAB-8

Quant Time: Mar 19 04:12:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	112752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	210234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	193367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	75482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	62570	58.554	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.100%
35) Dibromofluoromethane	7.728	113	70422	48.070	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.140%
50) Toluene-d8	10.191	98	239851	46.827	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.660%
62) 4-Bromofluorobenzene	12.489	95	66646	41.402	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.800%
Target Compounds						
20) Methylene Chloride	4.722	84	16262	8.846	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031824\
Data File : VY017698.D
Acq On : 18 Mar 2024 18:59
Operator : SY/MD
Sample : P1764-08
Misc : 4.54g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GRAB-8

Quant Time: Mar 19 04:12:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

