

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031824\
 Data File : VY017699.D
 Acq On : 18 Mar 2024 19:22
 Operator : SY/MD
 Sample : P1764-09
 Misc : 4.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GRAB-9

Quant Time: Mar 19 04:13:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	112145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	207653	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.496	117	190378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	73669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	60465	56.890	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.780%
35) Dibromofluoromethane	7.728	113	68849	47.581	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.160%
50) Toluene-d8	10.191	98	237385	46.921	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.840%
62) 4-Bromofluorobenzene	12.489	95	65875	41.431	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.860%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.728	84	13616	6.473	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031824\
Data File : VY017699.D
Acq On : 18 Mar 2024 19:22
Operator : SY/MD
Sample : P1764-09
Misc : 4.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GRAB-9

Quant Time: Mar 19 04:13:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

