

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031925\
 Data File : VY021565.D
 Acq On : 19 Mar 2025 12:40
 Operator : SY/MD
 Sample : Q1592-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OILY-DEBRIS-COMP

Quant Time: Mar 20 01:29:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	253134	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	466194	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	406514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	146155	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	146236	54.658	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.320%
35) Dibromofluoromethane	7.634	113	157160	51.286	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.580%
50) Toluene-d8	10.109	98	564936	48.690	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.380%
62) 4-Bromofluorobenzene	12.407	95	162892	41.273	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031925\
Data File : VY021565.D
Acq On : 19 Mar 2025 12:40
Operator : SY/MD
Sample : Q1592-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OILY-DEBRIS-COMP

Quant Time: Mar 20 01:29:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

