

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032025\
 Data File : VY021574.D
 Acq On : 20 Mar 2025 11:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 03/21/2025
 Supervised By :Mahesh Dadoda 03/21/2025

Quant Time: Mar 21 01:31:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	219959	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	337182	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	298070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	155271	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	123206	52.996	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 106.000%			
35) Dibromofluoromethane	7.634	113	119181	53.774	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 107.540%			
50) Toluene-d8	10.103	98	436168	51.976	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 103.960%			
62) 4-Bromofluorobenzene	12.408	95	153625	53.818	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 107.640%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.867	85	87379	43.289	ug/l	99
3) Chloromethane	2.068	50	123896	43.537	ug/l	99
4) Vinyl Chloride	2.202	62	149067	47.933	ug/l	99
5) Bromomethane	2.592	94	98811	43.962	ug/l	97
6) Chloroethane	2.732	64	105696	51.430	ug/l	95
7) Trichlorofluoromethane	3.056	101	232267	53.509	ug/l	100
8) Diethyl Ether	3.452	74	58873	48.941	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.811	101	118294	47.732	ug/l	99
10) Methyl Iodide	4.000	142	131060	52.025	ug/l	99
11) Tert butyl alcohol	4.860	59	44298	297.329	ug/l	99
12) 1,1-Dichloroethene	3.787	96	107582	47.614	ug/l	99
13) Acrolein	3.653	56	5463	46.199	ug/l	95
14) Allyl chloride	4.378	41	173427	47.496	ug/l	99
15) Acrylonitrile	5.055	53	141285	275.906	ug/l	99
16) Acetone	3.866	43	112533	221.929	ug/l	91
17) Carbon Disulfide	4.104	76	305583	43.155	ug/l	98
18) Methyl Acetate	4.385	43	79985	71.092	ug/l	99
19) Methyl tert-butyl Ether	5.110	73	296937	51.851	ug/l	98
20) Methylene Chloride	4.610	84	121289	45.568	ug/l	96
21) trans-1,2-Dichloroethene	5.110	96	120656	47.683	ug/l	94
22) Diisopropyl ether	6.018	45	384292	48.836	ug/l	97
23) Vinyl Acetate	5.957	43	1121418	249.707	ug/l	99
24) 1,1-Dichloroethane	5.909	63	225375	48.527	ug/l	98
25) 2-Butanone	6.884	43	181386	258.596	ug/l	98
26) 2,2-Dichloropropane	6.878	77	199701	46.732	ug/l	100
27) cis-1,2-Dichloroethene	6.890	96	140435	48.853	ug/l	98
28) Bromochloromethane	7.244	49	107010	54.959	ug/l	98
29) Tetrahydrofuran	7.256	42	121370	280.283	ug/l	99
30) Chloroform	7.421	83	238370	49.427	ug/l	99
31) Cyclohexane	7.701	56	190199	43.683	ug/l	99
32) 1,1,1-Trichloroethane	7.616	97	212459	48.114	ug/l	99
36) 1,1-Dichloropropene	7.835	75	164126	47.308	ug/l	98
37) Ethyl Acetate	6.982	43	83554	54.490	ug/l	99
38) Carbon Tetrachloride	7.817	117	194773	48.816	ug/l	97
39) Methylcyclohexane	9.109	83	200777	47.181	ug/l	99
40) Benzene	8.079	78	495197	48.455	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032025\
 Data File : VY021574.D
 Acq On : 20 Mar 2025 11:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 03/21/2025
 Supervised By :Mahesh Dadoda 03/21/2025

Quant Time: Mar 21 01:31:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.219	41	46502	55.696	ug/l	97
42) 1,2-Dichloroethane	8.158	62	143072	50.042	ug/l	98
43) Isopropyl Acetate	8.195	43	161556	54.274	ug/l #	90
44) Trichloroethene	8.865	130	128598	49.796	ug/l	100
45) 1,2-Dichloropropane	9.140	63	121269	49.858	ug/l	97
46) Dibromomethane	9.231	93	72294	53.418	ug/l	99
47) Bromodichloromethane	9.420	83	182499	50.778	ug/l	99
48) Methyl methacrylate	9.219	41	74817	54.736	ug/l	95
49) 1,4-Dioxane	9.225	88	17568	1322.530	ug/l	96
51) 4-Methyl-2-Pentanone	9.999	43	443231	288.587	ug/l	99
52) Toluene	10.170	92	329232	51.098	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	163401	51.604	ug/l	99
54) cis-1,3-Dichloropropene	9.853	75	189945	50.525	ug/l	99
55) 1,1,2-Trichloroethane	10.572	97	93743	52.952	ug/l	98
56) Ethyl methacrylate	10.438	69	129404	57.314	ug/l	98
57) 1,3-Dichloropropane	10.719	76	160374	52.268	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.707	63	287815	276.147	ug/l	99
59) 2-Hexanone	10.761	43	295385	290.862	ug/l	99
60) Dibromochloromethane	10.908	129	127078	52.360	ug/l	100
61) 1,2-Dibromoethane	11.011	107	87667	52.470	ug/l	100
64) Tetrachloroethene	10.646	164	137523	56.739	ug/l	99
65) Chlorobenzene	11.438	112	346955	49.292	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.517	131	125682	50.504	ug/l	99
67) Ethyl Benzene	11.517	91	611432	49.491	ug/l	98
68) m/p-Xylenes	11.627	106	469024	99.997	ug/l	100
69) o-Xylene	11.956	106	220832	51.085	ug/l	98
70) Styrene	11.969	104	377454	52.633	ug/l	99
71) Bromoform	12.133	173	75841	54.518	ug/l #	97
73) Isopropylbenzene	12.255	105	597306	48.835	ug/l	99
74) N-amyl acetate	12.072	43	140643	52.947	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.505	83	109695	50.823	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	81330m	51.980	ug/l	
77) Bromobenzene	12.529	156	139019	48.583	ug/l	99
78) n-propylbenzene	12.597	91	720756	48.522	ug/l	100
79) 2-Chlorotoluene	12.682	91	409969	48.268	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	488166	49.112	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	34651	50.345	ug/l	96
82) 4-Chlorotoluene	12.779	91	423488	48.149	ug/l	99
83) tert-Butylbenzene	12.999	119	442622	49.917	ug/l	98
84) 1,2,4-Trimethylbenzene	13.042	105	488589	49.508	ug/l	99
85) sec-Butylbenzene	13.176	105	646008	48.966	ug/l	99
86) p-Isopropyltoluene	13.291	119	545177	50.159	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	277496	48.719	ug/l	99
88) 1,4-Dichlorobenzene	13.371	146	266588	47.857	ug/l	99
89) n-Butylbenzene	13.615	91	509654	50.780	ug/l	99
90) Hexachloroethane	13.883	117	112588	48.556	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	244964	50.178	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.279	75	17141	53.016	ug/l	99
93) 1,2,4-Trichlorobenzene	14.919	180	142761	53.539	ug/l	97
94) Hexachlorobutadiene	15.023	225	87330	52.247	ug/l	99
95) Naphthalene	15.145	128	257284	56.217	ug/l	100
96) 1,2,3-Trichlorobenzene	15.334	180	123783	55.195	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032025\
Data File : VY021574.D
Acq On : 20 Mar 2025 11:09
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025

Quant Time: Mar 21 01:31:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 21995
Ion Ratio Lower Upper
168 100
99 53.3 46.0 69.0




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#2
Dichlorodifluoromethane
Concen: 43.289 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 85 Resp: 87379
Ion Ratio Lower Upper
85 100
87 30.9 15.7 47.0





#3
Chloromethane
Concen: 43.537 ug/l
RT: 2.068 min Scan# 51
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 50 Resp: 123890
Ion Ratio Lower Upper
50 100
52 32.1 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025




#4
Vinyl Chloride
Concen: 47.933 ug/l
RT: 2.202 min Scan# 79
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Tgt Ion: 62 Resp: 149067
Ion Ratio Lower Upper
62 100
64 32.0 26.2 39.2





#5
Bromomethane
Concen: 43.962 ug/l
RT: 2.592 min Scan# 145
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 94 Resp: 9881
Ion Ratio Lower Upper
94 100
96 92.2 76.3 114.5

Manual Integrations
APPROVED



Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025




#6
Chloroethane
Concen: 51.430 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09



Tgt Ion: 64 Resp: 105696
Ion Ratio Lower Upper
64 100
66 30.7 26.7 40.1









#7
Trichlorofluoromethane
Concen: 53.509 ug/l
RT: 3.056 min Scan# 21
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 23226
Ion Ratio Lower Upper
101 100
103 64.7 51.7 77.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#8
Diethyl Ether
Concen: 48.941 ug/l
RT: 3.452 min Scan# 284
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Tgt Ion: 74 Resp: 58873
Ion Ratio Lower Upper
74 100
45 96.9 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.732 ug/l
RT: 3.811 min Scan# 34
Delta R.T. -0.019 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 118294
Ion Ratio Lower Upper
101 100
85 47.2 37.1 55.7
151 78.2 62.1 93.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#10
Methyl Iodide
Concen: 52.025 ug/l
RT: 4.000 min Scan# 374
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion:142 Resp: 131060
Ion Ratio Lower Upper
142 100
127 47.9 39.1 58.7
141 14.4 11.8 17.6





#11
Tert butyl alcohol
Concen: 297.329 ug/l
RT: 4.860 min Scan# 517
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 59 Resp: 44298
Ion Ratio Lower Upper
59 100
57 10.5 8.2 12.4

Manual Integrations
APPROVED



Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025




#12
1,1-Dichloroethene
Concen: 47.614 ug/l
RT: 3.787 min Scan# 339
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Tgt Ion: 96 Resp: 107582
Ion Ratio Lower Upper
96 100
61 168.2 135.2 202.8
98 65.0 51.4 77.2





#13
Acrolein
Concen: 46.199 ug/l
RT: 3.653 min Scan# 317
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 546
Ion Ratio Lower Upper
56 100
55 74.6 63.4 95.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#14
Allyl chloride
Concen: 47.496 ug/l
RT: 4.378 min Scan# 436
Delta R.T. -0.018 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 41 Resp: 173427
Ion Ratio Lower Upper
41 100
39 73.2 59.3 88.9
76 38.5 29.8 44.8





#15
Acrylonitrile
Concen: 275.906 ug/l
RT: 5.055 min Scan# 54
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 141285
Ion Ratio Lower Upper
53 100
52 81.8 66.3 99.5
51 37.4 30.3 45.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#16
Acetone
Concen: 221.929 ug/l
RT: 3.866 min Scan# 352
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Tgt Ion: 43 Resp: 112533
Ion Ratio Lower Upper
43 100
58 32.4 22.1 33.1









#17
Carbon Disulfide
Concen: 43.155 ug/l
RT: 4.104 min Scan# 391
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 76 Resp: 30558
Ion Ratio Lower Upper
76 100
78 8.6 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#18
Methyl Acetate
Concen: 71.092 ug/l
RT: 4.385 min Scan# 437
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 43 Resp: 79985
Ion Ratio Lower Upper
43 100
74 24.9 19.4 29.0





#19
Methyl tert-butyl Ether
Concen: 51.851 ug/l
RT: 5.110 min Scan# 51
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 73 Resp: 29693
Ion Ratio Lower Upper
73 100
57 21.2 17.9 26.9




Manual Integrations
APPROVED

Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#20
Methylene Chloride
Concen: 45.568 ug/l
RT: 4.610 min Scan# 474
Delta R.T. -0.012 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 84 Resp: 121289
Ion Ratio Lower Upper
84 100
49 121.8 102.0 153.0
51 38.9 33.0 49.6
86 61.3 50.5 75.7





#21
trans-1,2-Dichloroethene
Concen: 47.683 ug/l
RT: 5.110 min Scan# 5110
Delta R.T. -0.018 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 120650
Ion Ratio Lower Upper
96 100
61 134.1 111.5 167.3
98 59.8 53.8 80.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#22
Diisopropyl ether
Concen: 48.836 ug/l
RT: 6.018 min Scan# 705
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 45 Resp: 384292
Ion Ratio Lower Upper
45 100
43 57.9 48.7 73.1
87 26.3 21.5 32.3
59 11.2 9.3 13.9





#23
 Vinyl Acetate
 Concen: 249.707 ug/l
 RT: 5.957 min Scan# 64
 Delta R.T. -0.012 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 1121418
 Ion Ratio Lower Upper
 43 100
 86 11.1 8.5 12.7

Manual Integrations
 APPROVED



Reviewed By :Amit Patel 03/21/2025
 Supervised By :Mahesh Dadoda 03/21/2025




#24
 1,1-Dichloroethane
 Concen: 48.527 ug/l
 RT: 5.909 min Scan# 687
 Delta R.T. -0.018 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09



Tgt Ion: 63 Resp: 225375
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.3 9.9
 100 4.4 2.1 6.5





#25

2-Butanone

Concen: 258.596 ug/l

RT: 6.884 min Scan# 84

Delta R.T. -0.012 min

Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 18138

Ion Ratio Lower Upper

43 100





72 25.9 21.4 32.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 03/21/2025

Supervised By :Mahesh Dadoda 03/21/2025



#26

2,2-Dichloropropane

Concen: 46.732 ug/l

RT: 6.878 min Scan# 846

Delta R.T. -0.012 min

Lab File: VY021574.D




Acq: 20 Mar 2025 11:09


Tgt Ion: 77 Resp: 199701

Ion Ratio Lower Upper

77 100

97 22.3 11.3 33.8









#27
 cis-1,2-Dichloroethene
 Concen: 48.853 ug/l
 RT: 6.890 min Scan# 849
 Delta R.T. -0.006 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Tgt Ion: 96 Resp: 140435
 Ion Ratio Lower Upper
 96 100
 61 146.9 0.0 300.8
 98 65.1 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Amit Patel 03/21/2025
 Supervised By :Mahesh Dadoda 03/21/2025



#28
 Bromochloromethane
 Concen: 54.959 ug/l
 RT: 7.244 min Scan# 906
 Delta R.T. -0.006 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09

Tgt Ion: 49 Resp: 107010
 Ion Ratio Lower Upper
 49 100
 129 2.3 0.0 4.2
 130 76.4 59.6 89.4





#29

Tetrahydrofuran

Concen: 280.283 ug/l

RT: 7.256 min Scan# 908

Delta R.T. -0.012 min

Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 121370





Ion	Ratio	Lower	Upper
42	100		
72	43.8	34.8	52.2
71	41.4	32.6	48.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 03/21/2025

Supervised By :Mahesh Dadoda 03/21/2025



#30

Chloroform

Concen: 49.427 ug/l

RT: 7.421 min Scan# 935




Delta R.T. -0.006 min


Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Tgt Ion: 83 Resp: 238370

Ion	Ratio	Lower	Upper
83	100		
85	66.0	53.1	79.7





#31

Cyclohexane

Concen: 43.683 ug/l

RT: 7.701 min Scan# 982

Delta R.T. -0.006 min

Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 190199

Ion Ratio Lower Upper

56 100

69 34.9 27.0 40.4





84 89.8 71.1 106.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 03/21/2025

Supervised By :Mahesh Dadoda 03/21/2025



#32

1,1,1-Trichloroethane

Concen: 48.114 ug/l

RT: 7.616 min Scan# 967

Delta R.T. -0.006 min

Lab File: VY021574.D

Acq: 20 Mar 2025 11:09




Tgt Ion: 97 Resp: 212459


Ion Ratio Lower Upper

97 100

99 65.9 51.7 77.5

61 44.7 35.7 53.5









#33
1,2-Dichloroethane-d4
Concen: 52.996 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 65 Resp: 123200
Ion Ratio Lower Upper
65 100
67 51.4 0.0 102.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.616 min Scan# 1131
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Tgt Ion:114 Resp: 337182
Ion Ratio Lower Upper
114 100
63 18.8 0.0 40.8
88 15.0 0.0 30.8





#35

Dibromofluoromethane

Concen: 53.774 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.006 min

Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 113 Resp: 11918





Ion	Ratio	Lower	Upper
113	100		
111	101.4	82.0	123.0
192	19.9	15.9	23.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 03/21/2025

Supervised By :Mahesh Dadoda 03/21/2025



#36

1,1-Dichloropropene

Concen: 47.308 ug/l

RT: 7.835 min Scan# 1003




Delta R.T. -0.006 min


Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Tgt Ion: 75 Resp: 164126

Ion	Ratio	Lower	Upper
75	100		
110	35.6	17.3	51.7
77	31.1	25.4	38.2





#37

Ethyl Acetate

Concen: 54.490 ug/l

RT: 6.982 min Scan# 80

Delta R.T. -0.006 min

Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 83554





Ion	Ratio	Lower	Upper
43	100		
61	13.7	11.4	17.0
70	10.3	8.2	12.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 03/21/2025

Supervised By :Mahesh Dadoda 03/21/2025



#38

Carbon Tetrachloride

Concen: 48.816 ug/l

RT: 7.817 min Scan# 1000




Delta R.T. -0.006 min


Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Tgt Ion:117 Resp: 194773


Ion	Ratio	Lower	Upper
117	100		
119	97.5	75.8	113.8
121	30.7	23.9	35.9





#39
Methylcyclohexane
Concen: 47.181 ug/l
RT: 9.109 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 20077
Ion Ratio Lower Upper
83 100
55 74.3 60.7 91.1
98 48.5 38.6 57.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#40
Benzene
Concen: 48.455 ug/l
RT: 8.079 min Scan# 1043
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Tgt Ion: 78 Resp: 495197
Ion Ratio Lower Upper
78 100
77 24.1 19.0 28.4









#41
Methacrylonitrile
Concen: 55.696 ug/l
RT: 7.219 min Scan# 902
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 41 Resp: 46502
Ion Ratio Lower Upper
41 100
39 47.2 35.0 52.6
67 74.9 61.2 91.8
52 32.5 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#42
1,2-Dichloroethane
Concen: 50.042 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 62 Resp: 143072
Ion Ratio Lower Upper
62 100
98 9.7 0.0 18.2





#43
Isopropyl Acetate
Concen: 54.274 ug/l
RT: 8.195 min Scan# 1062
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument : MSVOA_Y
ClientSampleId : VSTDCCC050



Tgt Ion: 43 Resp: 161550
Ion Ratio Lower Upper
43 100
61 21.4 23.4 35.0
87 13.3 10.5 15.7




Manual Integrations
APPROVED

Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#44
Trichloroethene
Concen: 49.796 ug/l
RT: 8.865 min Scan# 1172
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Tgt Ion:130 Resp: 128598
Ion Ratio Lower Upper
130 100
95 96.2 0.0 191.6





#45

1,2-Dichloropropane

Concen: 49.858 ug/l

RT: 9.140 min Scan# 121269

Delta R.T. -0.006 min

Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 121269

Ion Ratio Lower Upper

63 100





65 30.8 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 03/21/2025

Supervised By :Mahesh Dadoda 03/21/2025



#46

Dibromomethane

Concen: 53.418 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.006 min

Lab File: VY021574.D

Acq: 20 Mar 2025 11:09




Tgt Ion: 93 Resp: 72294


Ion Ratio Lower Upper

93 100

95 80.8 65.8 98.8


174 98.0 77.9 116.9





#47
Bromodichloromethane
Concen: 50.778 ug/l
RT: 9.420 min Scan# 1264
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 182499
Ion Ratio Lower Upper
83 100
85 64.9 51.1 76.7
127 8.3 6.7 10.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#48
Methyl methacrylate
Concen: 54.736 ug/l
RT: 9.219 min Scan# 1230
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 41 Resp: 74817
Ion Ratio Lower Upper
41 100
69 90.0 67.8 101.8
39 60.1 46.4 69.6





#49
1,4-Dioxane
Concen: 1322.530 ug/l
RT: 9.225 min Scan# 1231
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 88 Resp: 17568
Ion Ratio Lower Upper
88 100
43 37.1 25.0 37.4
58 73.5 58.4 87.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#50
Toluene-d8
Concen: 51.976 ug/l
RT: 10.103 min Scan# 1375
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Tgt Ion: 98 Resp: 436168
Ion Ratio Lower Upper
98 100
100 65.2 52.1 78.1









#51
4-Methyl-2-Pentanone
Concen: 288.587 ug/l
RT: 9.999 min Scan# 1358
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 44323
Ion Ratio Lower Upper
43 100
58 39.6 31.4 47.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#52
Toluene
Concen: 51.098 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Tgt Ion: 92 Resp: 329232
Ion Ratio Lower Upper
92 100
91 174.6 138.2 207.2









#53
t-1,3-Dichloropropene
Concen: 51.604 ug/l
RT: 10.390 min Scan# 1423
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 16340
Ion Ratio Lower Upper
75 100
77 31.7 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#54
cis-1,3-Dichloropropene
Concen: 50.525 ug/l
RT: 9.853 min Scan# 1334
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 75 Resp: 189945
Ion Ratio Lower Upper
75 100
77 31.1 24.6 37.0
39 48.3 39.4 59.2





#55
1,1,2-Trichloroethane
Concen: 52.952 ug/l
RT: 10.572 min Scan# 1452
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050






Tgt Ion: 97 Resp: 9374


Ion	Ratio	Lower	Upper
97	100		
83	86.9	70.0	105.0
85	57.7	44.3	66.5
99	65.8	51.4	77.0

Manual Integrations
APPROVED

Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025





#56
Ethyl methacrylate
Concen: 57.314 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09



Tgt Ion: 69 Resp: 129404

Ion	Ratio	Lower	Upper
69	100		
41	65.7	53.3	79.9
39	37.1	31.0	46.6





#57

1,3-Dichloropropane

Concen: 52.268 ug/l

RT: 10.719 min Scan# 1476

Delta R.T. 0.000 min

Lab File: VY021574.D

Acq: 20 Mar 2025 11:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 160374

Ion Ratio Lower Upper

76 100





78 32.0 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 03/21/2025

Supervised By :Mahesh Dadoda 03/21/2025



#58

2-Chloroethyl Vinyl ether

Concen: 276.147 ug/l

RT: 9.707 min Scan# 1310

Delta R.T. -0.006 min

Lab File: VY021574.D




Acq: 20 Mar 2025 11:09


Tgt Ion: 63 Resp: 287815

Ion Ratio Lower Upper

63 100


106 28.2 22.1 33.1





#59
2-Hexanone
Concen: 290.862 ug/l
RT: 10.761 min Scan# 1483
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 295385
Ion Ratio Lower Upper
43 100
58 55.2 27.3 81.8

Manual Integrations
APPROVED



Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025




#60
Dibromochloromethane
Concen: 52.360 ug/l
RT: 10.908 min Scan# 1507
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Tgt Ion:129 Resp: 127078
Ion Ratio Lower Upper
129 100
127 77.9 39.1 117.3





#61
1,2-Dibromoethane
Concen: 52.470 ug/l
RT: 11.011 min Scan# 1525
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 107 Resp: 8766
Ion Ratio Lower Upper
107 100
109 94.9 75.8 113.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#62
4-Bromofluorobenzene
Concen: 53.818 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 95 Resp: 153625
Ion Ratio Lower Upper
95 100
174 83.6 0.0 165.0
176 80.4 0.0 160.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 11590
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 298070
Ion Ratio Lower Upper
117 100
82 55.5 44.6 67.0
119 31.7 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025




#64
Tetrachloroethene
Concen: 56.739 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Tgt Ion:164 Resp: 137523
Ion Ratio Lower Upper
164 100
166 127.6 100.4 150.6
129 93.6 73.8 110.8
131 89.2 71.5 107.3





#65
Chlorobenzene
Concen: 49.292 ug/l
RT: 11.438 min Scan# 1594
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:112 Resp: 34695
Ion Ratio Lower Upper
112 100
114 31.1 26.1 39.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 50.504 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Tgt Ion:131 Resp: 125682
Ion Ratio Lower Upper
131 100
133 96.2 47.6 142.9
119 64.9 32.3 96.9





#67
Ethyl Benzene
Concen: 49.491 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 61143
Ion Ratio Lower Upper
91 100
106 31.6 24.6 36.8

Manual Integrations
APPROVED



Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025




#68
m/p-Xylenes
Concen: 99.997 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09



Tgt Ion:106 Resp: 469024
Ion Ratio Lower Upper
106 100
91 202.4 161.7 242.5









#69
o-Xylene
Concen: 51.085 ug/l
RT: 11.956 min Scan# 1679
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 220832
Ion Ratio Lower Upper
106 100
91 212.5 107.6 322.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#70
Styrene
Concen: 52.633 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion:104 Resp: 377454
Ion Ratio Lower Upper
104 100
78 49.8 40.7 61.1
103 54.5 44.1 66.1





#71
Bromoform
Concen: 54.518 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 7584
Ion Ratio Lower Upper
173 100
175 49.6 23.9 71.7
254 0.1 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Tgt Ion:152 Resp: 155271
Ion Ratio Lower Upper
152 100
115 57.7 29.0 87.0
150 173.6 0.0 347.2





#73
Isopropylbenzene
Concen: 48.835 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 105 Resp: 597300
Ion Ratio Lower Upper
105 100
120 26.7 13.1 39.3




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#74
N-ethyl acetate
Concen: 52.947 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 43 Resp: 140643
Ion Ratio Lower Upper
43 100
70 44.4 34.6 51.8
55 26.5 21.1 31.7
61 24.8 19.4 29.0





#75
 1,1,2,2-Tetrachloroethane
 Concen: 50.823 ug/l
 RT: 12.505 min Scan# 1769
 Delta R.T. -0.006 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 83 Resp: 10969
 Ion Ratio Lower Upper
 83 100
 131 10.8 5.3 16.1
 85 63.1 31.9 95.5

Manual Integrations
 APPROVED



Reviewed By :Amit Patel 03/21/2025
 Supervised By :Mahesh Dadoda 03/21/2025




#76
 1,2,3-Trichloropropane
 Concen: 51.980 ug/l m
 RT: 12.554 min Scan# 1777
 Delta R.T. -0.006 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09




Tgt Ion: 75 Resp: 81330
 Ion Ratio Lower Upper
 75 100
 77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 48.583 ug/l
RT: 12.529 min Scan# 1773
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 156 Resp: 139019
Ion Ratio Lower Upper
156 100
77 184.7 93.6 280.8
158 95.8 48.3 144.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#78
n-propylbenzene
Concen: 48.522 ug/l
RT: 12.597 min Scan# 1784
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 91 Resp: 720756
Ion Ratio Lower Upper
91 100
120 22.7 11.3 33.8





#79
2-Chlorotoluene
Concen: 48.268 ug/l
RT: 12.682 min Scan# 1798
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 409969
Ion Ratio Lower Upper
91 100
126 33.9 17.2 51.6

Manual Integrations
APPROVED



Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025




#80
1,3,5-Trimethylbenzene
Concen: 49.112 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09



Tgt Ion:105 Resp: 488166
Ion Ratio Lower Upper
105 100
120 48.8 24.6 73.6









#81
trans-1,4-Dichloro-2-butene
Concen: 50.345 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 3465
Ion Ratio Lower Upper
75 100
53 95.7 81.0 121.4
89 47.0 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#82
4-Chlorotoluene
Concen: 48.149 ug/l
RT: 12.779 min Scan# 1814
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 91 Resp: 423488
Ion Ratio Lower Upper
91 100
126 34.3 17.0 51.0





#83
 tert-Butylbenzene
 Concen: 49.917 ug/l
 RT: 12.999 min Scan# 1850
 Delta R.T. 0.000 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion:119 Resp: 442622
 Ion Ratio Lower Upper
 119 100
 91 64.5 33.2 99.6
 134 26.1 13.5 40.5

Manual Integrations
 APPROVED



Reviewed By :Amit Patel 03/21/2025
 Supervised By :Mahesh Dadoda 03/21/2025




#84
 1,2,4-Trimethylbenzene
 Concen: 49.508 ug/l
 RT: 13.042 min Scan# 1857
 Delta R.T. -0.006 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09



Tgt Ion:105 Resp: 488589
 Ion Ratio Lower Upper
 105 100
 120 45.0 22.7 68.1









#85
 sec-Butylbenzene
 Concen: 48.966 ug/l
 RT: 13.176 min Scan# 1879
 Delta R.T. 0.000 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Tgt Ion:105 Resp: 646008
 Ion Ratio Lower Upper
 105 100
 134 20.3 9.9 29.7




Manual Integrations
 APPROVED


Reviewed By :Amit Patel 03/21/2025
 Supervised By :Mahesh Dadoda 03/21/2025



#86
 p-Isopropyltoluene
 Concen: 50.159 ug/l
 RT: 13.291 min Scan# 1898
 Delta R.T. -0.006 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09


Tgt Ion:119 Resp: 545177
 Ion Ratio Lower Upper
 119 100
 134 26.2 13.1 39.1
 91 24.0 12.2 36.6





#87
 1,3-Dichlorobenzene
 Concen: 48.719 ug/l
 RT: 13.285 min Scan# 1897
 Delta R.T. -0.006 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion:146 Resp: 277490
 Ion Ratio Lower Upper
 146 100
 111 40.0 20.0 60.0
 148 62.9 31.8 95.3




Manual Integrations
 APPROVED


Reviewed By :Amit Patel 03/21/2025
 Supervised By :Mahesh Dadoda 03/21/2025



#88
 1,4-Dichlorobenzene
 Concen: 47.857 ug/l
 RT: 13.371 min Scan# 1911
 Delta R.T. 0.000 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09

Tgt Ion:146 Resp: 266588
 Ion Ratio Lower Upper
 146 100
 111 39.3 19.6 58.8
 148 65.5 32.3 96.8









#89
n-Butylbenzene
Concen: 50.780 ug/l
RT: 13.615 min Scan# 1951
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 509654
Ion Ratio Lower Upper
91 100
92 52.9 26.1 78.3
134 25.6 12.6 37.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#90
Hexachloroethane
Concen: 48.556 ug/l
RT: 13.883 min Scan# 1995
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Tgt Ion:117 Resp: 112588
Ion Ratio Lower Upper
117 100
201 75.5 37.4 112.2









#91
1,2-Dichlorobenzene
Concen: 50.178 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 244964
Ion Ratio Lower Upper
146 100
111 40.4 20.7 62.1
148 63.7 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 53.016 ug/l
RT: 14.279 min Scan# 2060
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09


Tgt Ion: 75 Resp: 17141
Ion Ratio Lower Upper
75 100
155 89.6 43.5 130.5
157 108.6 54.4 163.2





#93
 1,2,4-Trichlorobenzene
 Concen: 53.539 ug/l
 RT: 14.919 min Scan# 2165
 Delta R.T. -0.006 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion:180 Resp: 14276
 Ion Ratio Lower Upper
 180 100
 182 96.6 46.7 140.0
 145 31.4 15.6 46.8

Manual Integrations
 APPROVED



Reviewed By :Amit Patel 03/21/2025
 Supervised By :Mahesh Dadoda 03/21/2025




#94
 Hexachlorobutadiene
 Concen: 52.247 ug/l
 RT: 15.023 min Scan# 2182
 Delta R.T. -0.006 min
 Lab File: VY021574.D
 Acq: 20 Mar 2025 11:09




Tgt Ion:225 Resp: 87330
 Ion Ratio Lower Upper
 225 100
 223 62.7 31.4 94.2
 227 64.3 31.8 95.4





#95
Naphthalene
Concen: 56.217 ug/l
RT: 15.145 min Scan# 211
Delta R.T. -0.006 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:128 Resp: 257284
Ion Ratio Lower Upper
128 100
127 13.3 10.6 15.8
129 11.0 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :Amit Patel 03/21/2025
Supervised By :Mahesh Dadoda 03/21/2025



#96
1,2,3-Trichlorobenzene
Concen: 55.195 ug/l
RT: 15.334 min Scan# 2233
Delta R.T. 0.000 min
Lab File: VY021574.D
Acq: 20 Mar 2025 11:09



Tgt Ion:180 Resp: 123783
Ion Ratio Lower Upper
180 100
182 94.1 47.2 141.6
145 32.6 16.4 49.1

