

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032025\
 Data File : VY021575.D
 Acq On : 20 Mar 2025 12:13
 Operator : SY/MD
 Sample : VY0320SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0320SBL01

Quant Time: Mar 21 01:32:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	297160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	543041	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	482322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	191152	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	172571	54.945	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.880%
35) Dibromofluoromethane	7.634	113	180987	50.704	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.400%
50) Toluene-d8	10.109	98	649651	48.068	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.140%
62) 4-Bromofluorobenzene	12.408	95	202436	44.033	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.060%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032025\
Data File : VY021575.D
Acq On : 20 Mar 2025 12:13
Operator : SY/MD
Sample : VY0320SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBL01

Quant Time: Mar 21 01:32:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY021575.D
Acq: 20 Mar 2025 12:13




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBL01


Tgt Ion:168 Resp: 297160
Ion Ratio Lower Upper
168 100
99 57.6 46.0 69.0



#33
1,2-Dichloroethane-d4
Concen: 54.945 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. -0.006 min
Lab File: VY021575.D
Acq: 20 Mar 2025 12:13

Tgt Ion: 65 Resp: 172571
Ion Ratio Lower Upper
65 100
67 51.1 0.0 102.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.616 min Scan# 1132
Delta R.T. -0.006 min
Lab File: VY021575.D
Acq: 20 Mar 2025 12:13




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBL01


Tgt Ion:114 Resp: 543041
Ion Ratio Lower Upper
114 100
63 20.0 0.0 40.8
88 15.1 0.0 30.8



#35
Dibromofluoromethane
Concen: 50.704 ug/l
RT: 7.634 min Scan# 970
Delta R.T. -0.006 min
Lab File: VY021575.D
Acq: 20 Mar 2025 12:13

Tgt Ion:113 Resp: 180987
Ion Ratio Lower Upper
113 100
111 103.1 82.0 123.0
192 19.5 15.9 23.9









#50
Toluene-d8
Concen: 48.068 ug/l
RT: 10.109 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY021575.D
Acq: 20 Mar 2025 12:13




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBL01


Tgt Ion: 98 Resp: 649651
Ion Ratio Lower Upper
98 100
100 65.1 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 44.033 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.000 min
Lab File: VY021575.D
Acq: 20 Mar 2025 12:13

Tgt Ion: 95 Resp: 202436
Ion Ratio Lower Upper
95 100
174 83.3 0.0 165.0
176 80.3 0.0 160.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY021575.D
Acq: 20 Mar 2025 12:13

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBL01

Tgt Ion:117 Resp: 482322
Ion Ratio Lower Upper
117 100
82 53.6 44.6 67.0
119 32.1 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.006 min
Lab File: VY021575.D
Acq: 20 Mar 2025 12:13

Tgt Ion:152 Resp: 191152
Ion Ratio Lower Upper
152 100
115 57.7 29.0 87.0
150 155.9 0.0 347.2

