

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007952.D
 Acq On : 21 Mar 2022 14:15
 Operator : SY/MD
 Sample : N1875-13
 Misc : 6.80g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10D

Quant Time: Mar 22 03:39:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	251500	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	470437	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	422302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	161770	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	165257	57.630	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.260%
35) Dibromofluoromethane	7.716	113	149248	45.174	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.340%
50) Toluene-d8	10.179	98	572863	58.413	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	116.820%
62) 4-Bromofluorobenzene	12.483	95	189291	39.921	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.840%
Target Compounds						
16) Acetone	3.948	43	12985	17.069	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
Data File : VY007952.D
Acq On : 21 Mar 2022 14:15
Operator : SY/MD
Sample : N1875-13
Misc : 6.80g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10D

Quant Time: Mar 22 03:39:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

