

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007954.D
 Acq On : 21 Mar 2022 15:02
 Operator : SY/MD
 Sample : N1983-05
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 124019

Quant Time: Mar 21 16:05:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	243020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	450496	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	411980	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	179038	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	155034	55.951	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.900%
35) Dibromofluoromethane	7.716	113	140143	44.296	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.600%
50) Toluene-d8	10.179	98	550002	58.581	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	117.160%
62) 4-Bromofluorobenzene	12.483	95	194701	42.880	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.760%
Target Compounds						
16) Acetone	3.954	43	3650	4.965	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
Data File : VY007954.D
Acq On : 21 Mar 2022 15:02
Operator : SY/MD
Sample : N1983-05
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
124019

Quant Time: Mar 21 16:05:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

