

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007958.D
 Acq On : 21 Mar 2022 16:35
 Operator : SY/MD
 Sample : N1958-05RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C-5-4-5RE

Quant Time: Mar 22 03:40:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	52632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	93976	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	77471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	29544	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	28831	48.043	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.080%
35) Dibromofluoromethane	7.716	113	27288	41.346	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.700%
50) Toluene-d8	10.179	98	112769	57.468	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.940%
62) 4-Bromofluorobenzene	12.483	95	34010	35.906	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	71.820%
Target Compounds					Qvalue	
16) Acetone	3.942	43	2493	15.660	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
Data File : VY007958.D
Acq On : 21 Mar 2022 16:35
Operator : SY/MD
Sample : N1958-05RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C-5-4-5RE

Quant Time: Mar 22 03:40:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

