

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007969.D
 Acq On : 21 Mar 2022 22:00
 Operator : SY/MD
 Sample : VY0321SBL02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0321SBL02

Quant Time: Mar 22 03:42:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	224507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	423272	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	387791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	163662	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	147286	57.538	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.080%
35) Dibromofluoromethane	7.716	113	130529	43.911	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.820%
50) Toluene-d8	10.179	98	522251	59.271	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	118.540%
62) 4-Bromofluorobenzene	12.477	95	181176	42.467	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
Data File : VY007969.D
Acq On : 21 Mar 2022 22:00
Operator : SY/MD
Sample : VY0321SBL02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0321SBL02

Quant Time: Mar 22 03:42:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

