

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007972.D
 Acq On : 21 Mar 2022 23:33
 Operator : SY/MD
 Sample : N1864-08
 Misc : 4.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MB-B2-030922-5-6-DUP

Quant Time: Mar 22 03:43:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	18763	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	32834	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	24203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	8697	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	9871	46.141	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.280%
35) Dibromofluoromethane	7.710	113	9551	41.420	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.840%
50) Toluene-d8	10.179	98	37838	54.934	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.860%
62) 4-Bromofluorobenzene	12.477	95	10329	31.211	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	62.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
Data File : VY007972.D
Acq On : 21 Mar 2022 23:33
Operator : SY/MD
Sample : N1864-08
Misc : 4.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MB-B2-030922-5-6-DUP

Quant Time: Mar 22 03:43:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration

