

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007967.D
 Acq On : 21 Mar 2022 20:27
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Quant Time: Mar 22 01:29:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	196	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	1037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	2206	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1927	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	518	231.791	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	463.580%#
35) Dibromofluoromethane	7.710	113	47	6.454	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	12.900%#
50) Toluene-d8	10.179	98	1571	74.241	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	148.480%#
62) 4-Bromofluorobenzene	12.483	95	158317	15146.835	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	30293.680%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	84	67.511	ug/l #	43
3) Chloromethane	2.113	50	1106	787.683	ug/l #	65
4) Vinyl Chloride	2.260	62	317	168.906	ug/l #	43
5) Bromomethane	2.656	94	593	501.906	ug/l #	74
6) Chloroethane	2.723	64	26	16.221	ug/l #	43
7) Trichlorofluoromethane	3.125	101	88	24.488	ug/l #	30
8) Diethyl Ether	3.515	74	91	68.231	ug/l	86
10) Methyl Iodide	4.082	142	98	45.676	ug/l #	35
12) 1,1-Dichloroethene	3.857	96	83	38.999	ug/l #	1
13) Acrolein	3.662	56	23	146.677	ug/l #	82
14) Allyl chloride	4.473	41	111	31.008	ug/l #	22
15) Acrylonitrile	5.168	53	82	115.368	ug/l #	17
16) Acetone	3.948	43	1224	2064.604	ug/l	98
18) Methyl Acetate	4.473	43	271	210.625	ug/l #	56
19) Methyl tert-butyl Ether	5.198	73	66	9.886	ug/l #	52
20) Methylene Chloride	4.716	84	8383	4011.382	ug/l #	80
21) trans-1,2-Dichloroethene	5.241	96	22	9.233	ug/l #	1
22) Diisopropyl ether	6.125	45	79	10.258	ug/l #	33
23) Vinyl Acetate	6.058	43	466	96.607	ug/l #	75
24) 1,1-Dichloroethane	6.027	63	25	5.763	ug/l #	82
25) 2-Butanone	6.990	43	1604	1689.115	ug/l #	88
26) 2,2-Dichloropropane	7.155	77	20	4.974	ug/l #	55
27) cis-1,2-Dichloroethene	6.966	96	40	13.987	ug/l #	1
29) Tetrahydrofuran	7.338	42	436	716.568	ug/l #	49
30) Chloroform	7.503	83	435	97.017	ug/l	93
31) Cyclohexane	7.789	56	49	11.370	ug/l #	82
36) 1,1-Dichloropropene	7.905	75	60	5.555	ug/l #	42
37) Ethyl Acetate	7.082	43	209	31.710	ug/l #	72
40) Benzene	8.161	78	135	4.277	ug/l #	1
41) Methacrylonitrile	7.320	41	160	39.363	ug/l #	97
43) Isopropyl Acetate	8.283	43	42	3.305	ug/l #	56
47) Bromodichloromethane	9.508	83	195	16.882	ug/l #	74
48) Methyl methacrylate	9.289	41	79	13.976	ug/l #	34
51) 4-Methyl-2-Pentanone	10.069	43	360	53.775	ug/l	92
52) Toluene	10.234	92	220	10.853	ug/l	82
53) t-1,3-Dichloropropene	10.453	75	54	4.382	ug/l #	43

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007967.D
 Acq On : 21 Mar 2022 20:27
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Quant Time: Mar 22 01:29:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
54) cis-1,3-Dichloropropene	9.923	75	29	2.110	ug/l #	49
55) 1,1,2-Trichloroethane	10.569	97	46	6.622	ug/l #	15
57) 1,3-Dichloropropane	10.789	76	39	3.367	ug/l #	43
58) 2-Chloroethyl Vinyl ether	9.777	63	26	5.315	ug/l #	48
59) 2-Hexanone	10.819	43	17	3.551	ug/l #	1
60) Dibromochloromethane	10.996	129	23	2.765	ug/l #	11
61) 1,2-Dibromoethane	11.075	107	46	6.845	ug/l #	3
64) Tetrachloroethene	10.715	164	29	1.689	ug/l #	1
65) Chlorobenzene	11.520	112	217	4.106	ug/l #	42
67) Ethyl Benzene	11.593	91	441	4.685	ug/l #	79
68) m/p-Xylenes	11.697	106	338	9.205	ug/l	76
69) o-Xylene	12.032	106	161	4.533	ug/l	94
70) Styrene	12.044	104	321	5.375	ug/l #	86
73) Isopropylbenzene	12.331	105	458	2.700	ug/l #	48
75) 1,1,2,2-Tetrachloroethane	12.483	83	155	4.245	ug/l #	11
76) 1,2,3-Trichloropropane	12.483	75	84985	3094.164	ug/l #	100
77) Bromobenzene	12.618	156	110	2.838	ug/l #	59
78) n-propylbenzene	12.672	91	514	2.510	ug/l #	53
79) 2-Chlorotoluene	12.752	91	252	2.187	ug/l	81
80) 1,3,5-Trimethylbenzene	12.813	105	394	2.825	ug/l	69
81) trans-1,4-Dichloro-2-b...	12.483	75	84985	6498.995	ug/l #	7
82) 4-Chlorotoluene	12.849	91	371	3.099	ug/l	80
83) tert-Butylbenzene	13.069	119	244	1.955	ug/l #	74
84) 1,2,4-Trimethylbenzene	13.117	105	612	4.476	ug/l	78
85) sec-Butylbenzene	13.258	105	518	2.769	ug/l #	56
86) p-Isopropyltoluene	13.367	119	387	2.548	ug/l #	69
87) 1,3-Dichlorobenzene	13.367	146	300	3.963	ug/l #	74
88) 1,4-Dichlorobenzene	13.434	146	407	5.364	ug/l #	83
89) n-Butylbenzene	13.703	91	567	3.890	ug/l	87
91) 1,2-Dichlorobenzene	13.727	146	241	3.539	ug/l	95
93) 1,2,4-Trichlorobenzene	15.007	180	523	12.151	ug/l	75
95) Naphthalene	15.233	128	2042	21.263	ug/l	96
96) 1,2,3-Trichlorobenzene	15.428	180	482	12.696	ug/l	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007967.D
 Acq On : 21 Mar 2022 20:27
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Quant Time: Mar 22 01:29:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB

Tgt Ion:168 Resp: 196
Ion Ratio Lower Upper
168 100
99 0.0 46.9 70.3#




#2
Dichlorodifluoromethane
Concen: 67.511 ug/l
RT: 1.906 min Scan# 14
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 85 Resp: 84
Ion Ratio Lower Upper
85 100
87 0.0 16.1 48.2#



Abundance Scan 48 (2.113 min): VY007871.D\data.ms (-41)



#3

Chloromethane

Concen: 787.683 ug/l

RT: 2.113 min Scan# 41

Delta R.T. 0.000 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Instrument :

MSVOA_Y

ClientSampleId :

BFB


Tgt Ion: 50 Resp: 1106

Ion Ratio Lower Upper


50 100

52 50.2 24.6 37.0#


Abundance Scan 48 (2.113 min): VY007967.D\data.ms




Abundance



Abundance Scan 48 (2.113 min): VY007967.D\data.ms (-1)



Abundance Scan 71 (2.253 min): VY007871.D\data.ms (-64)



#4

Vinyl Chloride

Concen: 168.906 ug/l

RT: 2.260 min Scan# 72

Delta R.T. 0.006 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27


Tgt Ion: 62 Resp: 317

Ion Ratio Lower Upper


62 100

64 0.0 25.0 37.6#


Abundance Scan 72 (2.260 min): VY007967.D\data.ms




Abundance



Abundance Scan 72 (2.260 min): VY007967.D\data.ms (-22)









#5
Bromomethane
Concen: 501.906 ug/l
RT: 2.656 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 94 Resp: 593
Ion Ratio Lower Upper
94 100
96 66.9 73.5 110.3#



#6
Chloroethane
Concen: 16.221 ug/l
RT: 2.723 min Scan# 148
Delta R.T. -0.073 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 64 Resp: 26
Ion Ratio Lower Upper
64 100
66 0.0 25.4 38.0#









#7
Trichlorofluoromethane
Concen: 24.488 ug/l
RT: 3.125 min Scan# 214
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:101 Resp: 88
Ion Ratio Lower Upper
101 100
103 6.9 48.0 72.0#



#8
Diethyl Ether
Concen: 68.231 ug/l
RT: 3.515 min Scan# 278
Delta R.T. -0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 74 Resp: 91
Ion Ratio Lower Upper
74 100
45 95.6 55.1 165.5









#10
Methyl Iodide
Concen: 45.676 ug/l
RT: 4.082 min Scan# 31
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:142 Resp: 98
Ion Ratio Lower Upper
142 100
127 0.0 41.5 62.3#
141 0.0 11.9 17.9#



#12
1,1-Dichloroethene
Concen: 38.999 ug/l
RT: 3.857 min Scan# 334
Delta R.T. -0.018 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 96 Resp: 83
Ion Ratio Lower Upper
96 100
61 0.0 145.8 218.6#
98 0.0 52.5 78.7#









#13
Acrolein
Concen: 146.677 ug/l
RT: 3.662 min Scan# 302
Delta R.T. -0.067 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 56 Resp: 23
Ion Ratio Lower Upper
56 100
55 87.0 57.6 86.4#



#14
Allyl chloride
Concen: 31.008 ug/l
RT: 4.473 min Scan# 435
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 41 Resp: 111
Ion Ratio Lower Upper
41 100
39 0.0 58.8 88.2#
76 0.0 25.1 37.7#





#15

Acrylonitrile

Concen: 115.368 ug/l

RT: 5.168 min Scan# 54

Delta R.T. 0.006 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Instrument :

MSVOA_Y

ClientSampleId :

BFB





Tgt Ion: 53 Resp: 82

Ion Ratio Lower Upper

53 100

52 0.0 67.0 100.4#

51 0.0 29.8 44.6#



#16

Acetone

Concen: 2064.604 ug/l

RT: 3.948 min Scan# 349

Delta R.T. 0.006 min

Lab File: VY007967.D




Acq: 21 Mar 2022 20:27


Tgt Ion: 43 Resp: 1224

Ion Ratio Lower Upper

43 100

58 28.5 22.0 33.0









#18
Methyl Acetate
Concen: 210.625 ug/l
RT: 4.473 min Scan# 41
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 43 Resp: 271
Ion Ratio Lower Upper
43 100
74 0.0 16.6 24.8#



#19
Methyl tert-butyl Ether
Concen: 9.886 ug/l
RT: 5.198 min Scan# 554
Delta R.T. -0.024 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 73 Resp: 66
Ion Ratio Lower Upper
73 100
57 0.0 18.6 27.8#









#20
Methylene Chloride
Concen: 4011.382 ug/l
RT: 4.716 min Scan# 41
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Tgt Ion: 84 Resp: 8383




Ion	Ratio	Lower	Upper
84	100		
49	117.3	114.4	171.6
51	32.2	35.3	52.9#
86	52.3	53.7	80.5#




#21
trans-1,2-Dichloroethene
Concen: 9.233 ug/l
RT: 5.241 min Scan# 561
Delta R.T. 0.018 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 96 Resp: 22

Ion	Ratio	Lower	Upper
96	100		
61	0.0	120.5	180.7#
98	0.0	49.8	74.8#









#22
Diisopropyl ether
Concen: 10.258 ug/l
RT: 6.125 min Scan# 705
Delta R.T. 0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Tgt Ion: 45 Resp: 79




Ion	Ratio	Lower	Upper
45	100		
43	0.0	49.3	73.9#
87	0.0	19.2	28.8#
59	0.0	8.4	12.6#




#23
Vinyl Acetate
Concen: 96.607 ug/l
RT: 6.058 min Scan# 695
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 43 Resp: 466

Ion	Ratio	Lower	Upper
43	100		
86	0.0	7.1	10.7#





#24

1,1-Dichloroethane

Concen: 5.763 ug/l

RT: 6.027 min Scan# 61

Delta R.T. 0.012 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27





Instrument :

MSVOA_Y

ClientSampleId :

BFB

Tgt Ion: 63	Resp:	25
Ion	Ratio	Lower Upper
63	100	
98	0.0	3.6 10.8#
100	0.0	2.0 5.8#



#25

2-Butanone

Concen: 1689.115 ug/l




RT: 6.990 min Scan# 848


Delta R.T. 0.006 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Tgt Ion: 43	Resp:	1604
Ion	Ratio	Lower Upper
43	100	
72	27.8	17.7 26.5#





#26

2,2-Dichloropropane

Concen: 4.974 ug/l

RT: 7.155 min Scan# 81

Delta R.T. 0.177 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Instrument :

MSVOA_Y

ClientSampleId :





BFB

Tgt Ion: 77 Resp: 20

Ion Ratio Lower Upper

77 100

97 0.0 10.4 31.4#



#27

cis-1,2-Dichloroethene

Concen: 13.987 ug/l

RT: 6.966 min Scan# 844

Delta R.T. -0.018 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27




Tgt Ion: 96 Resp: 40


Ion Ratio Lower Upper

96 100

61 0.0 0.0 319.8

98 0.0 0.0 127.0









#29
Tetrahydrofuran
Concen: 716.568 ug/l
RT: 7.338 min Scan# 906
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:	42	Resp:	436
Ion	Ratio	Lower	Upper
42	100		
72	13.5	30.3	45.5#
71	0.0	29.0	43.4#



#30
Chloroform
Concen: 97.017 ug/l
RT: 7.503 min Scan# 932
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion:	83	Resp:	435
Ion	Ratio	Lower	Upper
83	100		
85	58.9	51.8	77.6









#31
Cyclohexane
Concen: 11.370 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 56 Resp: 49
Ion Ratio Lower Upper
56 100
69 0.0 26.7 40.1#
84 81.8 67.2 100.8



#33
1,2-Dichloroethane-d4
Concen: 231.791 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. 0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 65 Resp: 518
Ion Ratio Lower Upper
65 100
67 35.5 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Instrument :

MSVOA_Y

ClientSampleId :

BFB





Tgt Ion:114 Resp: 1037

Ion Ratio Lower Upper

114 100

63 0.0 0.0 43.2

88 15.5 0.0 31.8



#35

Dibromofluoromethane

Concen: 6.454 ug/l

RT: 7.710 min Scan# 966

Delta R.T. -0.006 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27




Tgt Ion:113 Resp: 47


Ion Ratio Lower Upper

113 100

111 314.9 82.6 124.0#

192 0.0 17.6 26.4#





#36

1,1-Dichloropropene

Concen: 5.555 ug/l

RT: 7.905 min Scan# 91

Delta R.T. -0.012 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Instrument :





MSVOA_Y

ClientSampleId :

BFB

Tgt Ion: 75 Resp: 60

Ion	Ratio	Lower	Upper
75	100		
110	0.0	17.1	51.2#
77	0.0	24.4	36.6#



#37

Ethyl Acetate

Concen: 31.710 ug/l

RT: 7.082 min Scan# 863




Delta R.T. 0.012 min


Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Tgt Ion: 43 Resp: 209

Ion	Ratio	Lower	Upper
43	100		
61	0.0	9.7	14.5#
70	0.0	7.1	10.7#









#40
Benzene
Concen: 4.277 ug/l
RT: 8.161 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 78 Resp: 135
Ion Ratio Lower Upper
78 100
77 90.7 20.4 30.6#



#41
Methacrylonitrile
Concen: 39.363 ug/l
RT: 7.320 min Scan# 902
Delta R.T. 0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 41 Resp: 160
Ion Ratio Lower Upper
41 100
39 113.1 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0









#43
Isopropyl Acetate
Concen: 3.305 ug/l
RT: 8.283 min Scan# 1060
Delta R.T. 0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:	43	Resp:	42
Ion	Ratio	Lower	Upper
43	100		
61	0.0	20.2	30.2#
87	0.0	8.9	13.3#



#47
Bromodichloromethane
Concen: 16.882 ug/l
RT: 9.508 min Scan# 1261
Delta R.T. 0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion:	83	Resp:	195
Ion	Ratio	Lower	Upper
83	100		
85	44.3	51.9	77.9#
127	0.0	6.6	10.0#





#48

Methyl methacrylate

Concen: 13.976 ug/l

RT: 9.289 min Scan# 1225

Delta R.T. 0.000 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Instrument :

MSVOA_Y

ClientSampleId :

BFB





Tgt Ion: 41 Resp: 79

Ion Ratio Lower Upper

41 100

69 0.0 58.2 87.2#

39 29.1 47.4 71.0#



#50

Toluene-d8

Concen: 74.241 ug/l

RT: 10.179 min Scan# 1371

Delta R.T. 0.000 min

Lab File: VY007967.D




Acq: 21 Mar 2022 20:27


Tgt Ion: 98 Resp: 1571

Ion Ratio Lower Upper

98 100

100 62.8 50.2 75.2









#51
4-Methyl-2-Pentanone
Concen: 53.775 ug/l
RT: 10.069 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 43 Resp: 360
Ion Ratio Lower Upper
43 100
58 30.3 28.1 42.1



#52
Toluene
Concen: 10.853 ug/l
RT: 10.234 min Scan# 1380
Delta R.T. -0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 92 Resp: 220
Ion Ratio Lower Upper
92 100
91 203.2 142.6 213.8









#53
t-1,3-Dichloropropene
Concen: 4.382 ug/l
RT: 10.453 min Scan# 1418
Delta R.T. -0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 75 Resp: 54
Ion Ratio Lower Upper
75 100
77 0.0 25.4 38.2#



#54
cis-1,3-Dichloropropene
Concen: 2.110 ug/l
RT: 9.923 min Scan# 1329
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 75 Resp: 29
Ion Ratio Lower Upper
75 100
77 0.0 25.6 38.4#
39 94.9 47.6 71.4#





#55

1,1,2-Trichloroethane

Concen: 6.622 ug/l

RT: 10.569 min Scan# 1447

Delta R.T. -0.073 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Instrument :

MSVOA_Y

ClientSampleId :

BFB

Tgt Ion: 97 Resp: 46


Ion Ratio Lower Upper

97 100




83 0.0 69.8 104.8#

85 0.0 44.3 66.5#

99 0.0 49.8 74.6#



Abundance



#57

1,3-Dichloropropane

Concen: 3.367 ug/l

RT: 10.789 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY007967.D


Acq: 21 Mar 2022 20:27

Tgt Ion: 76 Resp: 39



Ion Ratio Lower Upper


76 100

78 0.0 25.0 37.6#



Abundance









#58
2-Chloroethyl Vinyl ether
Concen: 5.315 ug/l
RT: 9.777 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 63 Resp: 26
Ion Ratio Lower Upper
63 100
106 0.0 21.5 32.3#



#59
2-Hexanone
Concen: 3.551 ug/l
RT: 10.819 min Scan# 1476
Delta R.T. -0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 43 Resp: 17
Ion Ratio Lower Upper
43 100
58 135.3 24.4 73.4#









#60
Dibromochloromethane
Concen: 2.765 ug/l
RT: 10.996 min Scan# 11
Delta R.T. 0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:129 Resp: 23
Ion Ratio Lower Upper
129 100
127 0.0 38.7 116.1#



#61
1,2-Dibromoethane
Concen: 6.845 ug/l
RT: 11.075 min Scan# 1518
Delta R.T. -0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion:107 Resp: 46
Ion Ratio Lower Upper
107 100
109 0.0 75.4 113.2#









#62
4-Bromofluorobenzene
Concen: 15146.835 ug/l
RT: 12.483 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 95 Resp: 158317
Ion Ratio Lower Upper
95 100
174 71.8 0.0 177.0
176 69.3 0.0 175.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.496 min Scan# 1587
Delta R.T. 0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 117 Resp: 2206
Ion Ratio Lower Upper
117 100
82 46.4 43.0 64.4
119 39.1 25.4 38.0#









#64
Tetrachloroethene
Concen: 1.689 ug/l
RT: 10.715 min Scan# 1459
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt	Ion:164	Resp:	29
Ion	Ratio	Lower	Upper
164	100		
166	0.0	100.4	150.6#
129	0.0	71.4	107.0#
131	0.0	70.8	106.2#



#65
Chlorobenzene
Concen: 4.106 ug/l
RT: 11.520 min Scan# 1591
Delta R.T. 0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion:112	Resp:	217	
Ion	Ratio	Lower	Upper
112	100		
114	0.0	25.8	38.6#









#67
Ethyl Benzene
Concen: 4.685 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 91 Resp: 441
Ion Ratio Lower Upper
91 100
106 19.2 24.5 36.7#



#68
m/p-Xylenes
Concen: 9.205 ug/l
RT: 11.697 min Scan# 1620
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 106 Resp: 338
Ion Ratio Lower Upper
106 100
91 246.2 166.6 249.8









#69
o-Xylene
Concen: 4.533 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:106 Resp: 161
Ion Ratio Lower Upper
106 100
91 230.4 110.3 331.0



#70
Styrene
Concen: 5.375 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion:104 Resp: 321
Ion Ratio Lower Upper
104 100
78 45.2 42.4 63.6
103 43.3 44.4 66.6#









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:152 Resp: 1927
Ion Ratio Lower Upper
152 100
115 62.4 28.2 84.7
150 158.0 0.0 347.6



#73
Isopropylbenzene
Concen: 2.700 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion:105 Resp: 458
Ion Ratio Lower Upper
105 100
120 0.0 13.4 40.2#









#75
1,1,2,2-Tetrachloroethane
Concen: 4.245 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.097 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 83 Resp: 155
Ion Ratio Lower Upper
83 100
131 118.1 5.6 16.8#
85 108.4 32.3 96.8#



#76
1,2,3-Trichloropropane
Concen: 3094.164 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.152 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 75 Resp: 84985
Ion Ratio Lower Upper
75 100
77 1.2 0.0 0.0#









#77
Bromobenzene
Concen: 2.838 ug/l
RT: 12.618 min Scan# 1770
Delta R.T. 0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:156 Resp: 110
Ion Ratio Lower Upper
156 100
77 222.7 95.5 286.4
158 18.2 48.0 144.0#



#78
n-propylbenzene
Concen: 2.510 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 91 Resp: 514
Ion Ratio Lower Upper
91 100
120 0.0 11.3 33.9#





#79

2-Chlorotoluene

Concen: 2.187 ug/l

RT: 12.752 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Instrument :

MSVOA_Y

ClientSampleId :





BFB

Tgt Ion: 91 Resp: 252

Ion Ratio Lower Upper

91 100

126 22.6 16.9 50.6



#80

1,3,5-Trimethylbenzene

Concen: 2.825 ug/l

RT: 12.813 min Scan# 1803

Delta R.T. 0.000 min

Lab File: VY007967.D




Acq: 21 Mar 2022 20:27


Tgt Ion:105 Resp: 394

Ion Ratio Lower Upper

105 100

120 27.7 24.4 73.4









#81
trans-1,4-Dichloro-2-butene
Concen: 6498.995 ug/l
RT: 12.483 min Scan# 11
Delta R.T. 0.104 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 75 Resp: 84985
Ion Ratio Lower Upper
75 100
53 0.0 85.3 127.9#
89 0.0 37.0 55.4#



#82
4-Chlorotoluene
Concen: 3.099 ug/l
RT: 12.849 min Scan# 1809
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 91 Resp: 371
Ion Ratio Lower Upper
91 100
126 45.6 17.0 51.0









#83
tert-Butylbenzene
Concen: 1.955 ug/l
RT: 13.069 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:119 Resp: 244
Ion Ratio Lower Upper
119 100
91 82.4 34.8 104.4
134 0.0 13.3 39.9#



#84
1,2,4-Trimethylbenzene
Concen: 4.476 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion:105 Resp: 612
Ion Ratio Lower Upper
105 100
120 30.1 22.3 66.9









#85
 sec-Butylbenzene
 Concen: 2.769 ug/l
 RT: 13.258 min Scan# 1876
 Delta R.T. 0.006 min
 Lab File: VY007967.D
 Acq: 21 Mar 2022 20:27




Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB


Tgt Ion:105 Resp: 518
 Ion Ratio Lower Upper
 105 100
 134 0.0 10.2 30.4#



#86
 p-Isopropyltoluene
 Concen: 2.548 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY007967.D
 Acq: 21 Mar 2022 20:27

Tgt Ion:119 Resp: 387
 Ion Ratio Lower Upper
 119 100
 134 19.4 13.2 39.5
 91 0.0 12.4 37.2#





#87

1,3-Dichlorobenzene

Concen: 3.963 ug/l

RT: 13.367 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27





Instrument :

MSVOA_Y

ClientSampleId :

BFB

Tgt Ion:	146	Resp:	300
Ion Ratio	Lower	Upper	
146	100		
111	0.0	19.5	58.5#
148	65.7	31.6	95.0



#88

1,4-Dichlorobenzene

Concen: 5.364 ug/l




RT: 13.434 min Scan# 1905


Delta R.T. -0.012 min

Lab File: VY007967.D

Acq: 21 Mar 2022 20:27

Tgt Ion:	146	Resp:	407
Ion Ratio	Lower	Upper	
146	100		
111	16.0	19.3	57.8#
148	60.7	32.6	97.8









#89
n-Butylbenzene
Concen: 3.890 ug/l
RT: 13.703 min Scan# 1948
Delta R.T. 0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion: 91	Resp:	567
Ion Ratio	Lower	Upper
91	100	
92	62.6	26.3 78.9
134	20.6	13.0 38.9



#91
1,2-Dichlorobenzene
Concen: 3.539 ug/l
RT: 13.727 min Scan# 1953
Delta R.T. -0.012 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion: 146	Resp:	241
Ion Ratio	Lower	Upper
146	100	
111	38.2	20.3 60.9
148	69.7	32.3 96.9









#93
1,2,4-Trichlorobenzene
Concen: 12.151 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27




Instrument :
MSVOA_Y
ClientSampleId :
BFB


Tgt Ion:	180	Resp:	523
Ion	Ratio	Lower	Upper
180	100		
182	68.5	47.3	141.9
145	18.5	14.6	43.8



#95
Naphthalene
Concen: 21.263 ug/l
RT: 15.233 min Scan# 2200
Delta R.T. -0.006 min
Lab File: VY007967.D
Acq: 21 Mar 2022 20:27

Tgt Ion:	128	Resp:	2042
Ion	Ratio	Lower	Upper
128	100		
127	11.8	11.0	16.4
129	9.4	8.4	12.6





#96
 1,2,3-Trichlorobenzene
 Concen: 12.696 ug/l
 RT: 15.428 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VY007967.D
 Acq: 21 Mar 2022 20:27

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Tgt Ion:	180	Resp:	482
Ion	Ratio	Lower	Upper
180	100		
182	102.7	47.3	141.9
145	29.5	15.3	45.9

