

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007980.D
 Acq On : 22 Mar 2022 02:40
 Operator : SY/MD
 Sample : N2031-03
 Misc : 6.67g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CMP-B1-031822-2-5

Quant Time: Mar 22 03:45:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 13:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	221837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	420287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	377305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	141702	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	150836	59.634	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.260%
35) Dibromofluoromethane	7.716	113	132426	44.865	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.740%
50) Toluene-d8	10.179	98	518716	59.290	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	118.580%
62) 4-Bromofluorobenzene	12.483	95	172535	40.729	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.460%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
Data File : VY007980.D
Acq On : 22 Mar 2022 02:40
Operator : SY/MD
Sample : N2031-03
Misc : 6.67g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CMP-B1-031822-2-5

Quant Time: Mar 22 03:45:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 13:55:34 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY007980.D
Acq: 22 Mar 2022 02:40




Instrument :
MSVOA_Y
ClientSampleId :
CMP-B1-031822-2-5


Tgt Ion:168 Resp: 221837
Ion Ratio Lower Upper
168 100
99 59.8 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 59.634 ug/l
RT: 8.142 min Scan# 1037
Delta R.T. 0.006 min
Lab File: VY007980.D
Acq: 22 Mar 2022 02:40

Tgt Ion: 65 Resp: 150836
Ion Ratio Lower Upper
65 100
67 52.6 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VY007980.D

Acq: 22 Mar 2022 02:40

Instrument :

MSVOA_Y

ClientSampleId :

CMP-B1-031822-2-5





Tgt Ion:114 Resp: 420287

Ion Ratio Lower Upper

114 100

63 19.7 0.0 43.2

88 16.1 0.0 31.8



#35

Dibromofluoromethane

Concen: 44.865 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY007980.D

Acq: 22 Mar 2022 02:40




Tgt Ion:113 Resp: 132426


Ion Ratio Lower Upper

113 100

111 102.4 82.6 124.0

192 16.5 17.6 26.4#









#50
Toluene-d8
Concen: 59.290 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY007980.D
Acq: 22 Mar 2022 02:40




Instrument :
MSVOA_Y
ClientSampleId :
CMP-B1-031822-2-5


Tgt Ion: 98 Resp: 518716
Ion Ratio Lower Upper
98 100
100 65.2 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 40.729 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY007980.D
Acq: 22 Mar 2022 02:40

Tgt Ion: 95 Resp: 172535
Ion Ratio Lower Upper
95 100
174 74.0 0.0 177.0
176 73.6 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 117
Delta R.T. -0.000 min
Lab File: VY007980.D
Acq: 22 Mar 2022 02:40

Instrument :
MSVOA_Y
ClientSampleId :
CMP-B1-031822-2-5

Tgt Ion:	117	Resp:	377305
Ion	Ratio	Lower	Upper
117	100		
82	55.8	43.0	64.4
119	32.1	25.4	38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY007980.D
Acq: 22 Mar 2022 02:40

Tgt Ion:	152	Resp:	141702
Ion	Ratio	Lower	Upper
152	100		
115	62.4	28.2	84.7
150	158.3	0.0	347.6

