



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032122\
 Data File : VY007942.D
 Acq On : 21 Mar 2022 08:55
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Title : SW846 8260
 Last Update : Thu Mar 17 13:55:34 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	18741	PASS
75	95	30	60	51.2	52821	PASS
95	95	100	100	100.0	103205	PASS
96	95	5	9	6.5	6669	PASS
173	174	0.00	2	0.7	541	PASS
174	95	50	100	72.6	74893	PASS
175	174	5	9	7.5	5617	PASS
176	174	95	101	98.9	74037	PASS
177	176	5	9	6.6	4912	PASS