



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032222\
 Data File : VY007989.D
 Acq On : 22 Mar 2022 08:58
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y031722S.M
 Title : SW846 8260
 Last Update : Thu Mar 17 13:55:34 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1737

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	16822	PASS
75	95	30	60	53.2	44973	PASS
95	95	100	100	100.0	84507	PASS
96	95	5	9	6.6	5536	PASS
173	174	0.00	2	1.0	589	PASS
174	95	50	100	71.1	60117	PASS
175	174	5	9	7.6	4556	PASS
176	174	95	101	96.3	57875	PASS
177	176	5	9	6.8	3963	PASS