

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032622\
 Data File : VY008036.D
 Acq On : 26 Mar 2022 11:23
 Operator : SY/MD
 Sample : VY0326SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0326SBL01

Quant Time: Mar 28 09:26:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	145652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	258886	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	239492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	104284	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	81569	55.818	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.640%
35) Dibromofluoromethane	7.716	113	81011	49.910	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.820%
50) Toluene-d8	10.179	98	313742	51.305	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.620%
62) 4-Bromofluorobenzene	12.483	95	105294	48.805	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032622\
Data File : VY008036.D
Acq On : 26 Mar 2022 11:23
Operator : SY/MD
Sample : VY0326SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0326SBL01

Quant Time: Mar 28 09:26:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

