

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032622\
 Data File : VY008043.D
 Acq On : 26 Mar 2022 14:05
 Operator : SY/MD
 Sample : N2095-02
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-02-5.0-6.0

Quant Time: Mar 28 09:28:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	122969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	223376	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	210019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	90653	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	71896	58.274	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.540%
35) Dibromofluoromethane	7.716	113	70367	50.244	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.480%
50) Toluene-d8	10.179	98	269531	51.082	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.160%
62) 4-Bromofluorobenzene	12.483	95	92277	49.570	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032622\
Data File : VY008043.D
Acq On : 26 Mar 2022 14:05
Operator : SY/MD
Sample : N2095-02
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-02-5.0-6.0

Quant Time: Mar 28 09:28:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

