

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032622\
 Data File : VY008051.D
 Acq On : 26 Mar 2022 17:13
 Operator : SY/MD
 Sample : N2095-15
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-15-5.0-6.0

Quant Time: Mar 28 09:30:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	32517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	55918	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	48056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	20177	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	18507	56.728	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.460%
35) Dibromofluoromethane	7.716	113	17239	49.171	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.340%
50) Toluene-d8	10.179	98	65026	49.230	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.460%
62) 4-Bromofluorobenzene	12.483	95	21405	45.933	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032622\
Data File : VY008051.D
Acq On : 26 Mar 2022 17:13
Operator : SY/MD
Sample : N2095-15
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-15-5.0-6.0

Quant Time: Mar 28 09:30:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

