

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032622\
 Data File : VY008054.D
 Acq On : 26 Mar 2022 18:23
 Operator : SY/MD
 Sample : N2095-14
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-14-4.0-5.0

Quant Time: Mar 28 09:31:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	21594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	38492	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	28902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	5773	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	13655	63.027	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.060%	
35) Dibromofluoromethane	7.716	113	12500	51.795	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.600%	
50) Toluene-d8	10.179	98	43379	47.710	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 95.420%	
62) 4-Bromofluorobenzene	12.483	95	9028	28.144	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 56.280%	
Target Compounds						
44) Trichloroethene	8.941	130	1652	5.268	ug/l	Qvalue 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032622\
Data File : VY008054.D
Acq On : 26 Mar 2022 18:23
Operator : SY/MD
Sample : N2095-14
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-14-4.0-5.0

Quant Time: Mar 28 09:31:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

