

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032625\
 Data File : VY021652.D
 Acq On : 26 Mar 2025 17:21
 Operator : SY/MD
 Sample : Q1636-15
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-4-VOC

Quant Time: Mar 27 01:33:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	3569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	5710	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	4433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	2927	77.594	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	155.180%
35) Dibromofluoromethane	7.634	113	2153	57.363	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.720%
50) Toluene-d8	10.109	98	6915	48.659	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.320%
62) 4-Bromofluorobenzene	12.414	95	1056	21.845	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	43.700%
Target Compounds						
52) Toluene	10.164	92	251	2.300	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032625\
Data File : VY021652.D
Acq On : 26 Mar 2025 17:21
Operator : SY/MD
Sample : Q1636-15
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-4-VOC

Quant Time: Mar 27 01:33:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

