

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032725\
 Data File : VY021675.D
 Acq On : 27 Mar 2025 18:32
 Operator : SY/MD
 Sample : Q1648-09
 Misc : 5.90g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-7

Quant Time: Mar 28 04:04:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	220283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	432276	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	403867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	163681	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	146770	61.495	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.980%	
35) Dibromofluoromethane	7.634	113	149722	53.394	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.780%	
50) Toluene-d8	10.109	98	533951	50.053	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.100%	
62) 4-Bromofluorobenzene	12.408	95	165171	45.775	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 91.560%	
Target Compounds						
					Qvalue	
64) Tetrachloroethene	10.646	164	14094	3.832	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032725\
Data File : VY021675.D
Acq On : 27 Mar 2025 18:32
Operator : SY/MD
Sample : Q1648-09
Misc : 5.90g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-7

Quant Time: Mar 28 04:04:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

