

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032725\
 Data File : VY021680.D
 Acq On : 27 Mar 2025 20:28
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 28 04:06:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.518	0.512	1.2	85	0.00
3 P	Chloromethane	0.729	0.753	-3.3	91	0.00
4 C	Vinyl Chloride	0.824	0.832	-1.0#	87	0.00
5 T	Bromomethane	0.562	0.581	-3.4	95	0.00
6 T	Chloroethane	0.539	0.557	-3.3	91	0.00
7 T	Trichlorofluoromethane	1.056	1.057	-0.1	87	0.00
8 T	Diethyl Ether	0.287	0.304	-5.9	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.569	-2.2	89	0.00
10 T	Methyl Iodide	0.611	0.679	-11.1	87	0.00
11 T	Tert butyl alcohol	0.035	0.038	-8.6	94	0.00
12 CM	1,1-Dichloroethene	0.524	0.541	-3.2#	88	0.00
13 T	Acrolein	0.064	0.072	-12.5	95	0.00
14 T	Allyl chloride	0.836	0.895	-7.1	90	0.00
15 T	Acrylonitrile	0.120	0.138	-15.0	96	0.00
16 T	Acetone	0.111	0.105	5.4	82	0.00
17 T	Carbon Disulfide	1.727	1.828	-5.8	90	0.00
18 T	Methyl Acetate	0.275	0.313	-13.8	99	0.00
19 T	Methyl tert-butyl Ether	1.306	1.485	-13.7	94	0.00
20 T	Methylene Chloride	0.626	0.647	-3.4	96	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.620	-7.5	92	0.00
22 T	Diisopropyl ether	1.754	2.038	-16.2	95	0.00
23 T	Vinyl Acetate	1.026	1.201	-17.1	94	0.00
24 P	1,1-Dichloroethane	1.049	1.148	-9.4	92	0.00
25 T	2-Butanone	0.158	0.173	-9.5	91	0.00
26 T	2,2-Dichloropropane	0.933	0.889	4.7	81	0.00
27 T	cis-1,2-Dichloroethene	0.649	0.722	-11.2	93	0.00
28 T	Bromochloromethane	0.442	0.524	-18.6	103	0.00
29 T	Tetrahydrofuran	0.102	0.122	-19.6	97	0.00
30 C	Chloroform	1.091	1.201	-10.1#	93	0.00
31 T	Cyclohexane	0.956	0.968	-1.3	88	0.00
32 T	1,1,1-Trichloroethane	0.986	1.044	-5.9	90	0.00
33 S	1,2-Dichloroethane-d4	0.542	0.620	-14.4	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.324	0.364	-12.3	93	0.00
36 T	1,1-Dichloropropene	0.494	0.515	-4.3	89	0.00
37 T	Ethyl Acetate	0.233	0.262	-12.4	97	0.00
38 T	Carbon Tetrachloride	0.566	0.580	-2.5	90	0.00
39 T	Methylcyclohexane	0.590	0.620	-5.1	87	0.00
40 TM	Benzene	1.479	1.591	-7.6	92	0.00
41 T	Methacrylonitrile	0.124	0.132	-6.5	84	-0.01
42 TM	1,2-Dichloroethane	0.417	0.453	-8.6	94	0.00
43 T	Isopropyl Acetate	0.442	0.493	-11.5	95	0.00
44 TM	Trichloroethene	0.377	0.390	-3.4	91	0.00
45 C	1,2-Dichloropropane	0.350	0.386	-10.3#	95	0.00
46 T	Dibromomethane	0.202	0.222	-9.9	97	0.00
47 T	Bromodichloromethane	0.523	0.573	-9.6	95	0.00
48 T	Methyl methacrylate	0.203	0.232	-14.3	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032725\
 Data File : VY021680.D
 Acq On : 27 Mar 2025 20:28
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 28 04:06:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	96	0.00
50 S	Toluene-d8	1.234	1.376	-11.5	92	0.00
51 T	4-Methyl-2-Pentanone	0.230	0.272	-18.3	98	0.00
52 CM	Toluene	0.923	1.028	-11.4#	94	0.00
53 T	t-1,3-Dichloropropene	0.465	0.511	-9.9	92	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.592	-8.0	91	0.00
55 T	1,1,2-Trichloroethane	0.260	0.291	-11.9	97	0.00
56 T	Ethyl methacrylate	0.332	0.400	-20.5	95	0.00
57 T	1,3-Dichloropropane	0.449	0.506	-12.7	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.162	0.199	-22.8	97	0.00
59 T	2-Hexanone	0.153	0.182	-19.0	97	0.00
60 T	Dibromochloromethane	0.355	0.398	-12.1	95	0.00
61 T	1,2-Dibromoethane	0.244	0.276	-13.1	95	0.00
62 S	4-Bromofluorobenzene	0.417	0.476	-14.1	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.455	0.451	0.9	90	0.00
65 PM	Chlorobenzene	1.146	1.189	-3.8	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.407	0.432	-6.1	94	0.00
67 C	Ethyl Benzene	1.965	2.111	-7.4#	92	0.00
68 T	m/p-Xylenes	0.753	0.807	-7.2	92	0.00
69 T	o-Xylene	0.695	0.748	-7.6	91	0.00
70 T	Styrene	1.158	1.309	-13.0	95	0.00
71 P	Bromoform	0.232	0.252	-8.6	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.714	3.757	-1.2	91	0.00
74 T	N-amyl acetate	0.872	0.932	-6.9	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.668	0.679	-1.6	96	0.00
76 T	1,2,3-Trichloropropane	0.496	0.521	-5.0	97	0.00
77 T	Bromobenzene	0.892	0.913	-2.4	95	0.00
78 T	n-propylbenzene	4.465	4.521	-1.3	91	0.00
79 T	2-Chlorotoluene	2.593	2.605	-0.5	91	0.00
80 T	1,3,5-Trimethylbenzene	3.035	3.086	-1.7	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.215	0.211	1.9	88	0.00
82 T	4-Chlorotoluene	2.704	2.715	-0.4	91	0.00
83 T	tert-Butylbenzene	2.703	2.703	0.0	90	0.00
84 T	1,2,4-Trimethylbenzene	3.000	3.108	-3.6	92	0.00
85 T	sec-Butylbenzene	3.951	3.954	-0.1	89	0.00
86 T	p-Isopropyltoluene	3.265	3.328	-1.9	90	0.00
87 T	1,3-Dichlorobenzene	1.757	1.754	0.2	93	0.00
88 T	1,4-Dichlorobenzene	1.736	1.704	1.8	91	0.00
89 T	n-Butylbenzene	3.020	3.019	0.0	88	0.00
90 T	Hexachloroethane	0.716	0.692	3.4	91	0.00
91 T	1,2-Dichlorobenzene	1.529	1.562	-2.2	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.103	0.108	-4.9	98	0.00
93 T	1,2,4-Trichlorobenzene	0.829	0.843	-1.7	89	0.00
94 T	Hexachlorobutadiene	0.518	0.502	3.1	89	0.00
95 T	Naphthalene	1.375	1.535	-11.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032725\
Data File : VY021680.D
Acq On : 27 Mar 2025 20:28
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 28 04:06:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.707	0.740	-4.7	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6