

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032825\
 Data File : VY021703.D
 Acq On : 28 Mar 2025 19:32
 Operator : SY/MD
 Sample : Q1664-17
 Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P001-BBDGA-006-01

Quant Time: Mar 28 23:48:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 28 02:30:29 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	216084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	410241	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	381513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	151297	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	133444	56.998	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.000%
35) Dibromofluoromethane	7.634	113	140354	52.742	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.480%
50) Toluene-d8	10.109	98	508912	50.269	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.540%
62) 4-Bromofluorobenzene	12.408	95	156593	45.729	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.460%
Target Compounds						Qvalue
20) Methylene Chloride	4.622	84	22927	6.406	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032825\
Data File : VY021703.D
Acq On : 28 Mar 2025 19:32
Operator : SY/MD
Sample : Q1664-17
Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P001-BBDGA-006-01

Quant Time: Mar 28 23:48:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032725S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 28 02:30:29 2025
Response via : Initial Calibration

