

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
 Data File : VY008082.D
 Acq On : 29 Mar 2022 12:58
 Operator : SY/MD
 Sample : N2142-01RE
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CW-1RE

Quant Time: Mar 30 03:58:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	74732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	134705	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.496	117	120861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	45622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	44123	58.847	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 117.700%	
35) Dibromofluoromethane	7.722	113	43147	51.088	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.180%	
50) Toluene-d8	10.185	98	160131	50.326	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.660%	
62) 4-Bromofluorobenzene	12.489	95	50844	45.292	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 90.580%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.129	105	4364	1.453	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
Data File : VY008082.D
Acq On : 29 Mar 2022 12:58
Operator : SY/MD
Sample : N2142-01RE
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CW-1RE

Quant Time: Mar 30 03:58:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

