

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
 Data File : VY008084.D
 Acq On : 29 Mar 2022 14:05
 Operator : SY/MD
 Sample : N2155-01
 Misc : 5.26g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 22L076-A

Quant Time: Mar 30 03:59:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	3505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	6561	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	5709	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	1592	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	2336	66.428	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.860%
35) Dibromofluoromethane	7.710	113	1211	29.439	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	58.880%
50) Toluene-d8	10.185	98	7529	48.581	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.160%
62) 4-Bromofluorobenzene	12.483	95	2289	41.864	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
Data File : VY008084.D
Acq On : 29 Mar 2022 14:05
Operator : SY/MD
Sample : N2155-01
Misc : 5.26g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
22L076-A

Quant Time: Mar 30 03:59:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

