

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
 Data File : VY008090.D
 Acq On : 29 Mar 2022 16:26
 Operator : SY/MD
 Sample : N2155-05
 Misc : 5.43g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 22L076-E

Quant Time: Mar 30 04:00:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	132567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	240344	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	227090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	99534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	76573	57.572	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.140%
35) Dibromofluoromethane	7.728	113	46605	30.928	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	61.860%
50) Toluene-d8	10.185	98	291882	51.413	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.820%
62) 4-Bromofluorobenzene	12.483	95	99683	49.768	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.540%
Target Compounds						
16) Acetone	3.954	43	5543	18.233	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
Data File : VY008090.D
Acq On : 29 Mar 2022 16:26
Operator : SY/MD
Sample : N2155-05
Misc : 5.43g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
22L076-E

Quant Time: Mar 30 04:00:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

