

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
 Data File : VY008093.D
 Acq On : 29 Mar 2022 17:36
 Operator : SY/MD
 Sample : N2125-03
 Misc : 5.31g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 164718

Quant Time: Mar 30 04:01:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	126587	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	229542	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	215895	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	93913	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	72058	56.736	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.480%
35) Dibromofluoromethane	7.722	113	72115	50.109	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.220%
50) Toluene-d8	10.185	98	275924	50.889	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.780%
62) 4-Bromofluorobenzene	12.483	95	94443	49.371	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
Data File : VY008093.D
Acq On : 29 Mar 2022 17:36
Operator : SY/MD
Sample : N2125-03
Misc : 5.31g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
164718

Quant Time: Mar 30 04:01:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

