

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
 Data File : VY008100.D
 Acq On : 29 Mar 2022 20:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda
 03/30/2022

Supervised By :Semsetun
 Yesilyurt
 03/30/2022

Quant Time: Mar 30 04:02:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	159273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	274858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	250057	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	123639	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.143	65	76569	47.916	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.840%
35) Dibromofluoromethane	7.716	113	85847	49.816	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.640%
50) Toluene-d8	10.179	98	316589	48.762	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.520%
62) 4-Bromofluorobenzene	12.483	95	111001	48.460	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.920%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.906	85	76254	53.110	ug/l	97
3) Chloromethane	2.113	50	76584	43.803	ug/l	95
4) Vinyl Chloride	2.253	62	99746	42.878	ug/l	97
5) Bromomethane	2.650	94	85386	44.788	ug/l	97
6) Chloroethane	2.796	64	67639	44.202	ug/l	98
7) Trichlorofluoromethane	3.131	101	137724	46.802	ug/l	92
8) Diethyl Ether	3.534	74	45181	47.411	ug/l	70
9) 1,1,2-Trichlorotrifluo...	3.906	101	87494	47.932	ug/l	92
10) Methyl Iodide	4.095	142	94846	44.672	ug/l	89
11) Tert butyl alcohol	4.954	59	34019	205.185	ug/l #	95
12) 1,1-Dichloroethene	3.875	96	81454	48.201	ug/l #	79
13) Acrolein	3.729	56	36025	266.954	ug/l	95
14) Allyl chloride	4.485	41	103122	46.381	ug/l #	90
15) Acrylonitrile	5.168	53	102277	225.319	ug/l	97
16) Acetone	3.948	43	70234	192.288	ug/l #	84
17) Carbon Disulfide	4.198	76	225199	47.031	ug/l	98
18) Methyl Acetate	4.479	43	41940	42.641	ug/l #	82
19) Methyl tert-butyl Ether	5.222	73	223569	46.553	ug/l	95
20) Methylene Chloride	4.723	84	92526	49.768	ug/l #	77
21) trans-1,2-Dichloroethene	5.222	96	95523	48.818	ug/l #	81
22) Diisopropyl ether	6.119	45	230147	48.113	ug/l #	91
23) Vinyl Acetate	6.064	43	733697	247.540	ug/l #	84
24) 1,1-Dichloroethane	6.021	63	152612	48.191	ug/l	97
25) 2-Butanone	6.990	43	117401	211.216	ug/l #	78
26) 2,2-Dichloropropane	6.984	77	139996	46.680	ug/l	91
27) cis-1,2-Dichloroethene	6.990	96	110828	48.865	ug/l	80
28) Bromochloromethane	7.338	49	40962	48.821	ug/l #	64
29) Tetrahydrofuran	7.344	42	79421	219.780	ug/l #	72
30) Chloroform	7.509	83	167983	48.989	ug/l	99
31) Cyclohexane	7.789	56	128162	42.907	ug/l #	81
32) 1,1,1-Trichloroethane	7.704	97	154596	48.795	ug/l	94
36) 1,1-Dichloropropene	7.917	75	124477	45.984	ug/l	95
37) Ethyl Acetate	7.076	43	54152	42.823	ug/l #	89
38) Carbon Tetrachloride	7.905	117	139086	47.148	ug/l	96
39) Methylcyclohexane	9.185	83	163050	44.625	ug/l #	83
40) Benzene	8.161	78	366787	47.369	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
 Data File : VY008100.D
 Acq On : 29 Mar 2022 20:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 30 04:02:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh
 Dadoda
 03/30/2022
 Supervised By :Semsettin
 Yesilyurt
 03/30/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.314	41	27610m	43.483	ug/l	
42) 1,2-Dichloroethane	8.240	62	100672	46.347	ug/l	89
43) Isopropyl Acetate	8.277	43	106194	42.778	ug/l #	82
44) Trichloroethene	8.941	130	106571	47.595	ug/l	100
45) 1,2-Dichloropropane	9.216	63	85526	46.867	ug/l	94
46) Dibromomethane	9.307	93	56087	46.743	ug/l	96
47) Bromodichloromethane	9.496	83	130081	47.409	ug/l	99
48) Methyl methacrylate	9.295	41	46410	42.210	ug/l #	71
49) 1,4-Dioxane	9.295	88	14180	888.961	ug/l #	76
51) 4-Methyl-2-Pentanone	10.069	43	271697	217.003	ug/l #	82
52) Toluene	10.246	92	243341	47.920	ug/l	95
53) t-1,3-Dichloropropene	10.465	75	131821	46.869	ug/l	100
54) cis-1,3-Dichloropropene	9.929	75	149087	46.826	ug/l #	82
55) 1,1,2-Trichloroethane	10.648	97	78499	47.379	ug/l	95
56) Ethyl methacrylate	10.508	69	98387	45.523	ug/l #	73
57) 1,3-Dichloropropane	10.789	76	127679	47.084	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.783	63	183908	205.376	ug/l #	87
59) 2-Hexanone	10.831	43	188094	214.691	ug/l	79
60) Dibromochloromethane	10.984	129	96527	47.903	ug/l	100
61) 1,2-Dibromoethane	11.087	107	77931	47.227	ug/l	98
64) Tetrachloroethene	10.721	164	88523	46.092	ug/l	97
65) Chlorobenzene	11.514	112	270920	48.051	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.593	131	101740	48.919	ug/l	97
67) Ethyl Benzene	11.593	91	466102	47.166	ug/l	97
68) m/p-Xylenes	11.703	106	373309	95.002	ug/l	90
69) o-Xylene	12.032	106	181105	48.136	ug/l	89
70) Styrene	12.044	104	303901	49.342	ug/l	94
71) Bromoform	12.209	173	60263	47.992	ug/l #	98
73) Isopropylbenzene	12.331	105	476383	46.680	ug/l	99
74) N-amyl acetate	12.142	43	95376	42.515	ug/l #	79
75) 1,1,2,2-Tetrachloroethane	12.581	83	88421	44.576	ug/l	99
76) 1,2,3-Trichloropropane	12.636	75	76878m	51.008	ug/l	
77) Bromobenzene	12.611	156	112902	48.223	ug/l	87
78) n-propylbenzene	12.672	91	556906	46.264	ug/l	96
79) 2-Chlorotoluene	12.758	91	307230	46.009	ug/l	93
80) 1,3,5-Trimethylbenzene	12.813	105	391659	46.847	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.380	75	28255	41.828	ug/l #	81
82) 4-Chlorotoluene	12.855	91	311397	45.448	ug/l	93
83) tert-Butylbenzene	13.075	119	352156	47.279	ug/l	92
84) 1,2,4-Trimethylbenzene	13.123	105	380699	46.760	ug/l	96
85) sec-Butylbenzene	13.258	105	510481	46.432	ug/l	97
86) p-Isopropyltoluene	13.373	119	429795	46.973	ug/l	96
87) 1,3-Dichlorobenzene	13.367	146	222583	47.953	ug/l	99
88) 1,4-Dichlorobenzene	13.447	146	220369	47.605	ug/l	98
89) n-Butylbenzene	13.697	91	384483	45.672	ug/l	96
90) Hexachloroethane	13.965	117	83011	46.824	ug/l	87
91) 1,2-Dichlorobenzene	13.739	146	196295	48.033	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.361	75	13668	43.172	ug/l	70
93) 1,2,4-Trichlorobenzene	15.007	180	115631	47.242	ug/l	98
94) Hexachlorobutadiene	15.111	225	61909	47.811	ug/l	99
95) Naphthalene	15.239	128	309193	57.149	ug/l	99
96) 1,2,3-Trichlorobenzene	15.428	180	102090	48.205	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032922\
Data File : VY008100.D
Acq On : 29 Mar 2022 20:19
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Quant Time: Mar 30 04:02:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
03/30/2022

Supervised By :Semsettin
Yesilyurt
03/30/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:168 Resp: 15927
Ion Ratio Lower Upper
168 100
99 52.2 46.9 70.3

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#2
Dichlorodifluoromethane
Concen: 53.110 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 85 Resp: 76254
Ion Ratio Lower Upper
85 100
87 33.7 16.1 48.2





#3
Chloromethane
Concen: 43.803 ug/l
RT: 2.113 min Scan# 41
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion: 50 Resp: 76584
Ion Ratio Lower Upper
50 100
52 33.6 24.6 37.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#4
Vinyl Chloride
Concen: 42.878 ug/l
RT: 2.253 min Scan# 71
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 62 Resp: 99746
Ion Ratio Lower Upper
62 100
64 32.9 25.0 37.6





#5
Bromomethane
Concen: 44.788 ug/l
RT: 2.650 min Scan# 11
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 94 Resp: 85380
Ion Ratio Lower Upper
94 100
96 94.3 73.5 110.3

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#6
Chloroethane
Concen: 44.202 ug/l
RT: 2.796 min Scan# 160
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 64 Resp: 67639
Ion Ratio Lower Upper
64 100
66 32.8 25.4 38.0





#7
Trichlorofluoromethane
Concen: 46.802 ug/l
RT: 3.131 min Scan# 214
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion:101 Resp: 137724
Ion Ratio Lower Upper
101 100
103 66.0 48.0 72.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022

Supervised By :Semsetin
Yesilyurt
03/30/2022






#8
Diethyl Ether
Concen: 47.411 ug/l
RT: 3.534 min Scan# 281
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion: 74 Resp: 45181
Ion Ratio Lower Upper
74 100
45 78.7 55.1 165.5




Abundance Scan 341 (3.899 min): VY008030.D\data.ms (-32



#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.932 ug/l
RT: 3.906 min Scan# 341
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion:	101	Resp:	87494
Ion	Ratio	Lower	Upper
101	100		
85	43.8	40.6	60.8
151	74.9	64.9	97.3



Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Manual Integrations

APPROVED

Reviewed By :Mahesh




Dadoda

03/30/2022

Supervised By :Semsetin


Yesilyurt


03/30/2022



#10
Methyl Iodide
Concen: 44.672 ug/l
RT: 4.095 min Scan# 373
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion:	142	Resp:	94846
Ion	Ratio	Lower	Upper
142	100		
127	42.8	41.5	62.3
141	13.9	11.9	17.9





#11

Tert butyl alcohol

Concen: 205.185 ug/l

RT: 4.954 min Scan# 513

Delta R.T. 0.006 min

Lab File: VY008100.D


Acq: 29 Mar 2022 20:19

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 59 Resp: 34019

Ion Ratio Lower Upper

59 100

57 9.8 6.4 9.6

Manual Integrations

APPROVED

Reviewed By :Mahesh



Dadoda

03/30/2022

Supervised By :Semsetun

Yesilyurt

03/30/2022



#12

1,1-Dichloroethene

Concen: 48.201 ug/l

RT: 3.875 min Scan# 337

Delta R.T. 0.006 min

Lab File: VY008100.D

Acq: 29 Mar 2022 20:19




Tgt Ion: 96 Resp: 81454


Ion Ratio Lower Upper

96 100

61 142.6 145.8 218.6#


98 65.3 52.5 78.7





#13
Acrolein
Concen: 266.954 ug/l
RT: 3.729 min Scan# 312
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 56 Resp: 36025
Ion Ratio Lower Upper
56 100
55 67.8 57.6 86.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#14
Allyl chloride
Concen: 46.381 ug/l
RT: 4.485 min Scan# 437
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 41 Resp: 103122
Ion Ratio Lower Upper
41 100
39 73.3 58.8 88.2
76 48.7 25.1 37.7#





#15
Acrylonitrile
Concen: 225.319 ug/l
RT: 5.168 min Scan# 547
Delta R.T. 0.013 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 53 Resp: 10227
Ion Ratio Lower Upper
53 100
52 81.0 67.0 100.4
51 36.3 29.8 44.6

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#16
Acetone
Concen: 192.288 ug/l
RT: 3.948 min Scan# 349
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 43 Resp: 70234
Ion Ratio Lower Upper
43 100
58 35.6 22.0 33.0#





#17
Carbon Disulfide
Concen: 47.031 ug/l
RT: 4.198 min Scan# 390
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 76 Resp: 225199
Ion Ratio Lower Upper
76 100
78 9.9 7.4 11.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#18
Methyl Acetate
Concen: 42.641 ug/l
RT: 4.479 min Scan# 436
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion: 43 Resp: 41940
Ion Ratio Lower Upper
43 100
74 29.3 16.6 24.8





#19

Methyl tert-butyl Ether

Concen: 46.553 ug/l

RT: 5.222 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY008100.D

Acq: 29 Mar 2022 20:19

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 73 Resp: 22356

Ion Ratio Lower Upper

73 100

57 20.6 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Mahesh





Dadoda

03/30/2022

Supervised By :Semsettin

Yesilyurt

03/30/2022



#20

Methylene Chloride

Concen: 49.768 ug/l

RT: 4.723 min Scan# 476

Delta R.T. 0.013 min

Lab File: VY008100.D

Acq: 29 Mar 2022 20:19

Tgt Ion: 84 Resp: 92526




Ion Ratio Lower Upper


84 100

49 101.7 114.4 171.6#

51 31.7 35.3 52.9#


86 65.8 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 48.818 ug/l
RT: 5.222 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 96 Resp: 9552
Ion Ratio Lower Upper
96 100
61 115.8 120.5 180.7
98 62.0 49.8 74.8

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#22
Diisopropyl ether
Concen: 48.113 ug/l
RT: 6.119 min Scan# 705
Delta R.T. 0.001 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion: 45 Resp: 230147
Ion Ratio Lower Upper
45 100
43 58.8 49.3 73.9
87 36.4 19.2 28.8#
59 13.0 8.4 12.6#






#23
 Vinyl Acetate
 Concen: 247.540 ug/l
 RT: 6.064 min Scan# 695
 Delta R.T. 0.007 min
 Lab File: VY008100.D
 Acq: 29 Mar 2022 20:19

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 43 Resp: 73369
 Ion Ratio Lower Upper
 43 100
 86 14.5 7.1 10.7

Manual Integrations
 APPROVED




Reviewed By :Mahesh
 Dadoda
 03/30/2022


Supervised By :Semsetin
 Yesilyurt
 03/30/2022



#24
 1,1-Dichloroethane
 Concen: 48.191 ug/l
 RT: 6.021 min Scan# 689
 Delta R.T. 0.006 min
 Lab File: VY008100.D
 Acq: 29 Mar 2022 20:19

Tgt Ion: 63 Resp: 152612
 Ion Ratio Lower Upper
 63 100
 98 7.9 3.6 10.8
 100 5.1 2.0 5.8





#25

2-Butanone

Concen: 211.216 ug/l

RT: 6.990 min Scan# 84

Delta R.T. 0.012 min

Lab File: VY008100.D

Acq: 29 Mar 2022 20:19

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 11740

Ion Ratio Lower Upper

43 100

72 32.6 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :Mahesh





Dadoda

03/30/2022

Supervised By :Semsetin

Yesilyurt

03/30/2022



#26

2,2-Dichloropropane

Concen: 46.680 ug/l

RT: 6.984 min Scan# 847

Delta R.T. 0.006 min

Lab File: VY008100.D




Acq: 29 Mar 2022 20:19


Tgt Ion: 77 Resp: 139996

Ion Ratio Lower Upper

77 100


97 25.0 10.4 31.4





#27
cis-1,2-Dichloroethene
Concen: 48.865 ug/l
RT: 6.990 min Scan# 847
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 96 Resp: 110828
Ion Ratio Lower Upper
96 100
61 124.5 0.0 319.8
98 65.1 0.0 127.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#28
Bromochloromethane
Concen: 48.821 ug/l
RT: 7.338 min Scan# 905
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 49 Resp: 40962
Ion Ratio Lower Upper
49 100
129 2.2 0.0 2.2#
130 99.1 55.4 83.0#





#29
Tetrahydrofuran
Concen: 219.780 ug/l
RT: 7.344 min Scan# 906
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 42 Resp: 7942
Ion Ratio Lower Upper
42 100
72 54.7 30.3 45.5
71 52.2 29.0 43.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#30
Chloroform
Concen: 48.989 ug/l
RT: 7.509 min Scan# 933
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 83 Resp: 167983
Ion Ratio Lower Upper
83 100
85 64.0 51.8 77.6





#31
Cyclohexane
Concen: 42.907 ug/l
RT: 7.789 min Scan# 911
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 56 Resp: 128162
Ion Ratio Lower Upper
56 100
69 36.7 26.7 40.1
84 106.1 67.2 100.8

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#32
1,1,1-Trichloroethane
Concen: 48.795 ug/l
RT: 7.704 min Scan# 965
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion: 97 Resp: 154596
Ion Ratio Lower Upper
97 100
99 64.5 50.7 76.1
61 38.8 37.6 56.4









#33
1,2-Dichloroethane-d4
Concen: 47.916 ug/l
RT: 8.143 min Scan# 1036
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 65 Resp: 76569
Ion Ratio Lower Upper
65 100
67 53.8 0.0 99.4




Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda
03/30/2022
Supervised By :Semsetin
Yesilyurt
03/30/2022



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.691 min Scan# 1127
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:114 Resp: 274858
Ion Ratio Lower Upper
114 100
63 16.3 0.0 43.2
88 14.6 0.0 31.8





#35
Dibromofluoromethane
Concen: 49.816 ug/l
RT: 7.716 min Scan# 90
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:113 Resp: 8584
Ion Ratio Lower Upper
113 100
111 102.7 82.6 124.0
192 18.2 17.6 26.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#36
1,1-Dichloropropene
Concen: 45.984 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 75 Resp: 124477
Ion Ratio Lower Upper
75 100
110 38.9 17.1 51.2
77 31.2 24.4 36.6





#37
Ethyl Acetate
Concen: 42.823 ug/l
RT: 7.076 min Scan# 861
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 43 Resp: 54152
Ion Ratio Lower Upper
43 100
61 16.6 9.7 14.5
70 12.9 7.1 10.7

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#38
Carbon Tetrachloride
Concen: 47.148 ug/l
RT: 7.905 min Scan# 998
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion:117 Resp: 139086
Ion Ratio Lower Upper
117 100
119 93.2 77.9 116.9
121 29.7 24.8 37.2





#39

Methylcyclohexane

Concen: 44.625 ug/l

RT: 9.185 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY008100.D

Acq: 29 Mar 2022 20:19

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 163050

Ion	Ratio	Lower	Upper
83	100		
55	62.2	66.5	99.7
98	51.6	37.9	56.9

Manual Integrations

APPROVED

Reviewed By :Mahesh





Dadoda

03/30/2022

Supervised By :Semsetin

Yesilyurt

03/30/2022



#40

Benzene

Concen: 47.369 ug/l

RT: 8.161 min Scan# 1040




Delta R.T. -0.000 min


Lab File: VY008100.D

Acq: 29 Mar 2022 20:19

Tgt Ion: 78 Resp: 366787

Ion	Ratio	Lower	Upper
78	100		
77	24.3	20.4	30.6






#41
Methacrylonitrile
Concen: 43.483 ug/l m
RT: 7.314 min Scan# 901
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 41 Resp: 27610
Ion Ratio Lower Upper
41 100
39 0.0 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#42
1,2-Dichloroethane
Concen: 46.347 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 62 Resp: 100672
Ion Ratio Lower Upper
62 100
98 11.2 0.0 14.8





#43
Isopropyl Acetate
Concen: 42.778 ug/l
RT: 8.277 min Scan# 1059
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 43 Resp: 106194
Ion Ratio Lower Upper
43 100
61 33.9 20.2 30.2
87 18.4 8.9 13.3

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#44
Trichloroethene
Concen: 47.595 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion:130 Resp: 106571
Ion Ratio Lower Upper
130 100
95 93.8 0.0 188.2





#45

1,2-Dichloropropane

Concen: 46.867 ug/l

RT: 9.216 min Scan# 1213

Delta R.T. 0.001 min

Lab File: VY008100.D

Acq: 29 Mar 2022 20:19

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 85520

Ion Ratio Lower Upper

63 100

65 31.8 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :Mahesh





Dadoda

03/30/2022

Supervised By :Semsetin

Yesilyurt

03/30/2022



#46

Dibromomethane

Concen: 46.743 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. 0.006 min

Lab File: VY008100.D

Acq: 29 Mar 2022 20:19




Tgt Ion: 93 Resp: 56087


Ion Ratio Lower Upper

93 100

95 83.9 64.3 96.5


174 99.3 83.4 125.0





#47
Bromodichloromethane
Concen: 47.409 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 83 Resp: 13008
Ion Ratio Lower Upper
83 100
85 64.2 51.9 77.9
127 8.4 6.6 10.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#48
Methyl methacrylate
Concen: 42.210 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 41 Resp: 46410
Ion Ratio Lower Upper
41 100
69 112.2 58.2 87.2#
39 64.7 47.4 71.0





#49
1,4-Dioxane
Concen: 888.961 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 88 Resp: 14180
Ion Ratio Lower Upper
88 100
43 25.0 32.8 49.2
58 55.5 59.8 89.6

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#50
Toluene-d8
Concen: 48.762 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 98 Resp: 316589
Ion Ratio Lower Upper
98 100
100 65.9 50.2 75.2





#51
4-Methyl-2-Pentanone
Concen: 217.003 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 43 Resp: 271697
Ion Ratio Lower Upper
43 100
58 45.4 28.1 42.1

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#52
Toluene
Concen: 47.920 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 92 Resp: 243341
Ion Ratio Lower Upper
92 100
91 171.2 142.6 213.8





#53
t-1,3-Dichloropropene
Concen: 46.869 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 13182
Ion Ratio Lower Upper
75 100
77 31.6 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda
03/30/2022
Supervised By :Semsettin
Yesilyurt
03/30/2022



#54
cis-1,3-Dichloropropene
Concen: 46.826 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 75 Resp: 149087
Ion Ratio Lower Upper
75 100
77 32.9 25.6 38.4
39 39.1 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 47.379 ug/l
RT: 10.648 min Scan# 1447
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 97 Resp: 7849

Ion	Ratio	Lower	Upper
97	100		
83	79.5	69.8	104.8
85	53.1	44.3	66.5
99	60.2	49.8	74.6

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
03/30/2022




Supervised By :Semsetin
Yesilyurt
03/30/2022




#56
Ethyl methacrylate
Concen: 45.523 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion: 69 Resp: 98387

Ion	Ratio	Lower	Upper
69	100		
41	53.4	63.4	95.0#
39	30.9	35.8	53.8#





#57

1,3-Dichloropropane

Concen: 47.084 ug/l

RT: 10.789 min Scan# 1471

Delta R.T. 0.001 min

Lab File: VY008100.D

Acq: 29 Mar 2022 20:19

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 127679

Ion Ratio Lower Upper

76 100

78 33.0 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Mahesh





Dadoda

03/30/2022

Supervised By :Semsetin

Yesilyurt

03/30/2022



#58

2-Chloroethyl Vinyl ether

Concen: 205.376 ug/l

RT: 9.783 min Scan# 1306

Delta R.T. 0.001 min

Lab File: VY008100.D




Acq: 29 Mar 2022 20:19


Tgt Ion: 63 Resp: 183908

Ion Ratio Lower Upper

63 100


106 33.5 21.5 32.3#





#59
2-Hexanone
Concen: 214.691 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 43 Resp: 188094
Ion Ratio Lower Upper
43 100
58 63.5 24.4 73.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#60
Dibromochloromethane
Concen: 47.903 ug/l
RT: 10.984 min Scan# 1503
Delta R.T. 0.001 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:129 Resp: 96527
Ion Ratio Lower Upper
129 100
127 77.4 38.7 116.1





#61
1,2-Dibromoethane
Concen: 47.227 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion:107 Resp: 7793
Ion Ratio Lower Upper
107 100
109 95.8 75.4 113.2


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
03/30/2022



Supervised By :Semsettin
Yesilyurt
03/30/2022




#62
4-Bromofluorobenzene
Concen: 48.460 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19




Tgt Ion: 95 Resp: 111001
Ion Ratio Lower Upper
95 100
174 87.5 0.0 177.0
176 84.2 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 11
Delta R.T. 0.001 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:117 Resp: 25005
Ion Ratio Lower Upper
117 100
82 50.2 43.0 64.4
119 31.5 25.4 38.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#64
Tetrachloroethene
Concen: 46.092 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:164 Resp: 88523
Ion Ratio Lower Upper
164 100
166 126.3 100.4 150.6
129 95.7 71.4 107.0
131 90.3 70.8 106.2





#65
Chlorobenzene
Concen: 48.051 ug/l
RT: 11.514 min Scan# 1590
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:112 Resp: 270920
Ion Ratio Lower Upper
112 100
114 31.8 25.8 38.6

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 48.919 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion:131 Resp: 101740
Ion Ratio Lower Upper
131 100
133 96.2 47.9 143.6
119 66.0 30.8 92.3









#67
Ethyl Benzene
Concen: 47.166 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 46610
Ion Ratio Lower Upper
91 100
106 32.5 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda
03/30/2022
Supervised By :Semsetin
Yesilyurt
03/30/2022



#68
m/p-Xylenes
Concen: 95.002 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:106 Resp: 373309
Ion Ratio Lower Upper
106 100
91 193.0 166.6 249.8





#69
o-Xylene
Concen: 48.136 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:106 Resp: 18110
Ion Ratio Lower Upper
106 100
91 203.6 110.3 331.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#70
Styrene
Concen: 49.342 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:104 Resp: 303901
Ion Ratio Lower Upper
104 100
78 44.9 42.4 63.6
103 54.6 44.4 66.6





#71
Bromoform
Concen: 47.992 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:173 Resp: 60263
Ion Ratio Lower Upper
173 100
175 48.1 24.6 74.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:152 Resp: 123639
Ion Ratio Lower Upper
152 100
115 55.3 28.2 84.7
150 175.1 0.0 347.6





#73
Isopropylbenzene
Concen: 46.680 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:105 Resp: 47638
Ion Ratio Lower Upper
105 100
120 27.4 13.4 40.2

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#74
N-ethyl acetate
Concen: 42.515 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 43 Resp: 95376
Ion Ratio Lower Upper
43 100
70 54.7 28.7 43.1#
55 25.9 17.3 25.9
61 28.7 17.0 25.6#





#75
1,1,2,2-Tetrachloroethane
Concen: 44.576 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 83 Resp: 8842
Ion Ratio Lower Upper
83 100
131 11.2 5.6 16.8
85 65.6 32.3 96.8

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#76
1,2,3-Trichloropropane
Concen: 51.008 ug/l m
RT: 12.636 min Scan# 1774
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 75 Resp: 76878
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 48.223 ug/l
RT: 12.611 min Scan# 1769
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:156 Resp: 112902
Ion Ratio Lower Upper
156 100
77 163.2 95.5 286.4
158 97.7 48.0 144.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#78
n-propylbenzene
Concen: 46.264 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion: 91 Resp: 556906
Ion Ratio Lower Upper
91 100
120 24.6 11.3 33.9









#79
2-Chlorotoluene
Concen: 46.009 ug/l
RT: 12.758 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 307230
Ion Ratio Lower Upper
91 100
126 37.4 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda
03/30/2022
Supervised By :Semsetin
Yesilyurt
03/30/2022



#80
1,3,5-Trimethylbenzene
Concen: 46.847 ug/l
RT: 12.813 min Scan# 1803
Delta R.T. 0.001 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:105 Resp: 391659
Ion Ratio Lower Upper
105 100
120 50.6 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 41.828 ug/l
RT: 12.380 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 75 Resp: 2825
Ion Ratio Lower Upper
75 100
53 82.2 85.3 127.9
89 52.1 37.0 55.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#82
4-Chlorotoluene
Concen: 45.448 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 91 Resp: 311397
Ion Ratio Lower Upper
91 100
126 37.9 17.0 51.0





#83
tert-Butylbenzene
Concen: 47.279 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:119 Resp: 352150
Ion Ratio Lower Upper
119 100
91 61.5 34.8 104.4
134 27.7 13.3 39.9

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#84
1,2,4-Trimethylbenzene
Concen: 46.760 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion:105 Resp: 380699
Ion Ratio Lower Upper
105 100
120 47.4 22.3 66.9





#85
sec-Butylbenzene
Concen: 46.432 ug/l
RT: 13.258 min Scan# 1875
Delta R.T. 0.007 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19





Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:105 Resp: 51048
Ion Ratio Lower Upper
105 100
134 21.6 10.2 30.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#86
p-Isopropyltoluene
Concen: 46.973 ug/l
RT: 13.373 min Scan# 1895
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:119 Resp: 429795
Ion Ratio Lower Upper
119 100
134 27.7 13.2 39.5
91 22.2 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 47.953 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:146 Resp: 222583
Ion Ratio Lower Upper
146 100
111 37.9 19.5 58.5
148 63.6 31.6 95.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsetin
Yesilyurt
03/30/2022



#88
1,4-Dichlorobenzene
Concen: 47.605 ug/l
RT: 13.447 min Scan# 1907
Delta R.T. 0.001 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:146 Resp: 220369
Ion Ratio Lower Upper
146 100
111 37.6 19.3 57.8
148 64.0 32.6 97.8





#89
n-Butylbenzene
Concen: 45.672 ug/l
RT: 13.697 min Scan# 1948
Delta R.T. 0.001 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 91 Resp: 38448
Ion Ratio Lower Upper
91 100
92 54.3 26.3 78.9
134 28.9 13.0 38.9

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#90
Hexachloroethane
Concen: 46.824 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion:117 Resp: 83011
Ion Ratio Lower Upper
117 100
201 73.7 43.0 128.8





#91
1,2-Dichlorobenzene
Concen: 48.033 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:146 Resp: 196295
Ion Ratio Lower Upper
146 100
111 39.3 20.3 60.9
148 64.1 32.3 96.9

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
03/30/2022


Supervised By :Semsettin
Yesilyurt
03/30/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 43.172 ug/l
RT: 14.361 min Scan# 2057
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19


Tgt Ion: 75 Resp: 13668
Ion Ratio Lower Upper
75 100
155 100.4 38.0 113.9
157 126.0 47.8 143.3





#93
1,2,4-Trichlorobenzene
Concen: 47.242 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:180 Resp: 11563
Ion Ratio Lower Upper
180 100
182 96.2 47.3 141.9
145 31.8 14.6 43.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
03/30/2022



Supervised By :Semsettin
Yesilyurt
03/30/2022




#94
Hexachlorobutadiene
Concen: 47.811 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19




Tgt Ion:225 Resp: 61909
Ion Ratio Lower Upper
225 100
223 62.4 30.8 92.3
227 63.2 31.6 94.7





#95
Naphthalene
Concen: 57.149 ug/l
RT: 15.239 min Scan# 2201
Delta R.T. -0.000 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:128 Resp: 30919
Ion Ratio Lower Upper
128 100
127 12.9 11.0 16.4
129 10.9 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
03/30/2022

Supervised By :Semsetin
Yesilyurt
03/30/2022



#96
1,2,3-Trichlorobenzene
Concen: 48.205 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.006 min
Lab File: VY008100.D
Acq: 29 Mar 2022 20:19

Tgt Ion:180 Resp: 102090
Ion Ratio Lower Upper
180 100
182 95.1 47.3 141.9
145 33.7 15.3 45.9

