

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033022\
 Data File : VY008110.D
 Acq On : 30 Mar 2022 14:55
 Operator : SY/MD
 Sample : N2155-03RE
 Misc : 4.92g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 22L076-CRE

Quant Time: Mar 31 17:21:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	6885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	12477	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	11479	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4378	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	4122	59.672	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.340%
35) Dibromofluoromethane	7.716	113	1846	23.598	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	47.200%#
50) Toluene-d8	10.179	98	15234	51.689	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.380%
62) 4-Bromofluorobenzene	12.483	95	4772	45.894	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033022\
Data File : VY008110.D
Acq On : 30 Mar 2022 14:55
Operator : SY/MD
Sample : N2155-03RE
Misc : 4.92g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
22L076-CRE

Quant Time: Mar 31 17:21:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

