

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033022\
 Data File : VY008112.D
 Acq On : 30 Mar 2022 15:41
 Operator : SY/MD
 Sample : N2155-07RE
 Misc : 5.23g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 22L076-GRE

Quant Time: Apr 11 11:13:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 15:45:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	122061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	222164	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	209857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	96668	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	71285	58.209	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 116.420%	
35) Dibromofluoromethane	7.722	113	32846	23.581	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 47.160%#	
50) Toluene-d8	10.179	98	270324	51.512	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.020%	
62) 4-Bromofluorobenzene	12.483	95	95052	51.340	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.680%	
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	130765	467.156	ug/l #	86
25) 2-Butanone	6.984	43	5625	13.205	ug/l #	82
52) Toluene	10.246	92	34793	8.477	ug/l	97
68) m/p-Xylenes	11.697	106	7316	2.218	ug/l	99
69) o-Xylene	12.032	106	3443	1.090	ug/l	91
84) 1,2,4-Trimethylbenzene	13.123	105	8766	1.377	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033022\
Data File : VY008112.D
Acq On : 30 Mar 2022 15:41
Operator : SY/MD
Sample : N2155-07RE
Misc : 5.23g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
22L076-GRE

Quant Time: Apr 11 11:13:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 15:45:35 2022
Response via : Initial Calibration

