

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033022\
 Data File : VY008122.D
 Acq On : 30 Mar 2022 19:59
 Operator : SY/MD
 Sample : N2176-02
 Misc : 5.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TT-156-IDWSO-20220329-2

Quant Time: Mar 31 05:03:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	124954	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	228490	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	208132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	83348	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	70672	56.372	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.740%
35) Dibromofluoromethane	7.716	113	72037	50.285	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.580%
50) Toluene-d8	10.179	98	274260	50.815	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.620%
62) 4-Bromofluorobenzene	12.483	95	88368	46.408	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033022\
Data File : VY008122.D
Acq On : 30 Mar 2022 19:59
Operator : SY/MD
Sample : N2176-02
Misc : 5.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TT-156-IDWSO-20220329-2

Quant Time: Mar 31 05:03:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

